



Received: 7 February 2023

Revised: 25 April 2023

Accepted: 28 April 2023

Blood microsampling technologies: Innovations and applications in 2022

Manchu Umarani Thangavelu¹  | Bert Wouters¹ | Alida Kindt¹  |
Irwin K. M. Reiss² | Thomas Hankemeier¹

¹Metabolomics and Analytics Centre, Leiden University, Leiden, The Netherlands

²Department of Neonatal and Pediatric Intensive Care, Division of Neonatology, Erasmus MC, Rotterdam, The Netherlands

Correspondence

Prof. dr. Thomas Hankemeier, Metabolomics and Analytics Centre, Leiden University, Einsteinweg 55, 2333 CC Leiden, The Netherlands.

Email: hankemeier@lacdr.leidenuniv.nl

Funding information

Medical Delta, Grant/Award Number: MOMETA; European Regional Development Fund, Grant/Award Number: Fieldlab MXL KvW-00267; Nederlandse Organisatie voor Wetenschappelijk Onderzoek, Grant/Award Number: Exposome-NL 024.004.017

Abstract

With the development of highly sensitive bioanalytical techniques, the volume of samples necessary for accurate analysis has reduced. Microsampling, the process of obtaining small amounts of blood, has thus gained popularity as it offers minimal-invasiveness, reduced logistical costs and biohazard risks while simultaneously showing increased sample stability and a potential for the decentralization of the approach and at-home self-sampling. Although the benefits of microsampling have been recognised, its adoption in clinical practice has been slow. Several microsampling technologies and devices are currently available and employed in research studies for various biomedical applications. This review provides an overview of the state-of-the-art in microsampling technology with a focus on the latest developments and advancements in the field of microsampling. Research published in the year 2022, including studies (i) developing strategies for the quantitation of analytes in microsamples and (ii) bridging and comparing the interchangeability between matrices and choice of technology for a given application, is reviewed to assess the advantages, challenges and limitations of the current state of microsampling. Successful implementation of microsampling in routine clinical care requires continued efforts for standardization and harmonization. Microsampling has been shown to facilitate data-rich studies and a patient-centric approach to healthcare and is foreseen to play a central role in the future digital revolution of healthcare through continuous monitoring to improve the quality of life.

Abbreviations: 3Rs, Replace, reduce and refine; 4OHCP, 4-hydroxycyclophosphamide; BRP, Biological reference preparation; CE-MS, Capillary electrophoresis-mass spectrometry; CERA, Continuous erythropoietin receptor activator; CMS, Capillary microsampling; CP, Cyclophosphamide; DBS, Dried blood spot; DPS, Dried plasma spot; DSS, Dried serum spot; EDTA, Ethylenediaminetetraacetic acid; EPO, Erythropoietin; ERA, Erythropoietin receptor agonist; Hct, Haematocrit; HRMS, High-resolution MS; IL-1Ra, Interleukin-1 receptor antagonist; IL-1 β , Interleukin-1 beta; IMS, Immunosuppressants; LC, Liquid chromatography; mAb, Monoclonal antibody; MRPL, Minimum required performance level; MS, Mass spectrometry; MSW, Microsampling Wing; NESP, Novel erythropoiesis stimulating protein; NIR, Near infrared; NPS, New psychoactive substances; PD, Pharmacodynamic; pDBS, Patterned DBS; PEth, Phosphatidylethanol; PG, Prostaglandin; PK, Pharmacokinetic; PSI, Paper spray ionization; RBC, Red blood cell; SPE, Solid-phase extraction; TD, Toxicodynamic; TDM, Therapeutic drug monitoring; Tepa, N,N,N'-triethylenephosphoramidate; THC, Tetrahydrocannabinol; Thiotepa, Triethylenethiophosphoramidate; TK, Toxicokinetic; TKI, Tyrosine kinase inhibitor; TMS, Tube microsampling; UHPLC, Ultra-high-performance LC; UPLC, Ultra-performance LC; UV/VIS, Ultra-violet/visible; VAMS, Volumetric absorptive microsampling®; VTM, Volumetric tip microsampling; WADA, World Anti-Doping Agency.

This is an open access article under the terms of the [Creative Commons Attribution](https://creativecommons.org/licenses/by/4.0/) License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.

© 2023 The Authors. *Analytical Science Advances* published by Wiley-VCH GmbH.

KEYWORDS

dried blood spot, forensic toxicology, microsampling, therapeutic drug monitoring, volumetric absorptive microsampling

1 | INTRODUCTION

Although invasive intravascular access has been the gold standard for blood sampling for decades and remains one of the most prevalent medical procedures in healthcare, its use is limited by several drawbacks.¹ The process of acquiring intravascular access with a hypodermic needle for blood sample collection (typically > 1 mL) requires a qualified phlebotomist and a sterile clinical environment.² This invasive and centralised approach is often associated with inconveniences such as frequent clinical visits, discomfort, anxiety, pain and phobia, which can lead to reduced patient compliance.³ Poor vascular puncture practices can result in haemoconcentration or haemolysis, rendering the samples unsuitable for analysis and subjecting patients to the inconvenience of a second blood draw. In addition, they expose healthcare personnel to the risk of sharp injuries and bloodborne pathogens. Occasionally, they may also lead to complications such as hematoma, infection, nerve damage and iatrogenic anaemia, potentially causing physical, psychological and economic distress.^{3–6} Consequently, medical procedures may be delayed or disregarded, and participation in clinical research may be reduced.⁷ Furthermore, these samples necessitate time-consuming, resource-intensive and expensive sample collection and preparation protocols to minimise pre-analytical variability.⁸ About 75% of samples require centrifugation to obtain plasma or serum for analysis, which increases economic costs and is a substantial factor for delays in a laboratory workflow.⁹ Wet blood samples require cold-chain storage and shipping to prevent sample degradation and bacterial growth.¹⁰

Blood microsampling, the process of capturing small volumes of capillary blood (typically <100 µL), often in a minimally-invasive manner, presents a viable alternative to vascular puncture and is a significant contributor in the revolutionization of human healthcare towards preventive, participative and personalised approaches in disease management. The minimal invasiveness of a low-volume blood draw provides benefits over conventional blood sampling by enabling amenable self-sampling without the need for trained personnel or hospital visits, which can advance primary healthcare in rural and remote geographical areas,¹¹ and improve patient recruitment and retention, thereby increasing statistical power in clinical studies.¹⁰ Reduced blood volumes enhance the feasibility of sampling volume-limited candidates and serial sampling, which can facilitate longitudinal investigations with fewer participants.¹² In addition, dried microsamples eliminate the requirement for centrifugation and aliquoting, may be transported at ambient temperatures without compromising analyte stability, can facilitate direct processing and are compatible with automated workflows. Dried matrices also render most pathogens inactive, reducing the biohazard risk generally associated with the transfer of clinical samples.¹³ The enhanced feasibility of sample processing, storage and

shipment significantly reduce the turnaround time of blood testing, hence promoting shorter time-to-diagnosis.¹⁴

The complexity of the blood matrix often necessitates elaborate sample clean-up procedures such as protein precipitation, liquid–liquid extraction and solid-phase extraction (SPE) to reduce matrix effects and enrich analytes to the desired detection limit.¹⁵ Since microsamples comprise only a few hundred microliters, highly sensitive bioanalytical assays are required for analysis in addition to improved sample clean-up, particularly for low-abundant target analytes. In addition to the conventional sample clean-up procedures, microsample preparation has been enhanced by the development of new techniques such as porous polymeric thin-film extraction,¹⁶ micro-SPE with pipette tips and spin columns¹⁷ and three-phase electroextraction, which is capable of achieving simultaneous clean-up and high enrichment of 20 µL microsamples in a short amount of time.¹⁸ Automation of miniaturised extraction techniques and direct hyphenation to analytical instrumentation minimises sample loss and alleviates the sample-preparation bottlenecks in volume-limited samples.¹⁹ The most extensively employed quantitative analytical technique for microsample analysis is liquid chromatography (LC) separation coupled with mass spectrometry (MS) detection due to its ability to detect analytes in pg/mL concentrations. Other separation techniques such as capillary electrophoresis-mass spectrometry (CE-MS) provide a potent analytical tool for the efficient profiling of polar and charged analytes in small sample volumes via separation by an electric field.²⁰ Furthermore, paper-based ionization techniques such as paper spray ionization (PSI) for MS have enabled the direct analysis of a broad spectrum of analytes in dried microsamples without the requirement of any prior sample pretreatment.²¹ These sophisticated approaches have driven the implementation of microsampling in various biomedical domains; newborn and metabolic screening,^{22,23} biomarker research,^{21,24,25} pharmacokinetic (PK) and pharmacodynamic (PD) studies,^{26,27} therapeutic drug monitoring (TDM),^{28–33} forensic toxicology,^{34–36} sports anti-doping,^{37–39} metabolomics^{22,40,41} and proteomics.^{24,42,43} Since 70% of medical decisions are governed by diagnostic blood tests,⁴⁴ several medical technology companies are investing in the development of innovative at-home blood microsampling devices to make blood work more accessible and convenient. Being an integral part of the paradigm shift towards patient-centric healthcare systems, the microsampling devices market is projected to reach a value of US\$ 2.4 Bn by 2030 at a compound growth annual rate of ~7%.⁴⁵

This review provides an extensive overview of the state-of-the-art in blood microsampling, with a focus on recent significant developments, technological advancements and applications, published in the year 2022. It includes a discussion on the working principles, advantages and challenges of various microsampling technologies and associated commercially-available devices. The key developments in




FIGURE 1 Graphical summary of the benefits of microsampling and 2022 publication statistics of microsampling technologies and applications.

microsampling, which include (i) development and validation of bio-analytical methods for the quantification of analytes, (ii) validation of microsampling by comparison with conventional sampling with emphasis on interchangeability between matrices and subsequent determination of matrix conversion factors if needed and (iii) evaluation and comparison of the performance of two or more microsampling devices to determine the optimum for a given application, are summarised and reviewed. Figure 1 provides a visual summary of the benefits of microsampling and statistics of microsampling technology and applications in 2022.

2 | MICROSAMPLING TECHNOLOGIES AND INNOVATIONS

Microsamples can be collected either as dried or liquid matrix microsamples. Existing technologies for collection of the former include the dried matrix spot, dried blood spheroid and volumetric tip microsampling. The latter can be collected either using a capillary-based or tube-based technology. The following sections provide a comprehensive overview of the microsampling technologies, including commercially-available devices and recent innovative concepts for microsampling.

2.1 | Dried matrix microsampling

An overview of commercially-available dried matrix microsampling devices can be found in Table 1.

2.1.1 | Dried matrix spot: Dried blood spot, dried plasma spot and dried serum spot

The dried blood spot (DBS) is the oldest microsampling technique and is well established in developed countries for newborn screening of more than 50 inborn genetic and metabolic disorders, while gaining popularity in various other biomedical applications.^{46–48} DBS samples are capillary blood droplets – from a lancet prick on the finger, heel or toe – dripped onto specially manufactured filter paper comprising a sheet of thick cellulose cardstock affixed to an envelope for sample identification and handling. The saturated filter paper is dried at ambient temperature for a minimum of 4 h and shipped to a laboratory for analysis, where fixed-diameter discs are punched out from the filter paper to provide volumetric defined measurements, and blood components are extracted by rehydrating the filter matrix with an elution buffer.¹⁵

Despite its simple blood collection process, DBS use in biomedical applications is limited by multiparametric sources of variability that

TABLE 1 Overview of commercially-available dried matrix microsampling devices.

Technology	Classification	Sub-classification	Device	Company	Design	Volume	Samples	Advantage	Disadvantage
Dried Matrix Spot	Dried Blood Spot	Non-Volumetric Dried Blood Spot	Conventional Filter Cards (Whatman 903)	Cytiva (Mascachusetts, United States)	rectangular filter paper held by a sturdy card or plastic frame to provide rigidity and stability during handling and processing	~80 μ L	5	cost effective; easy to handle; compatibility with automated lab processing	potential volume bias; easy contamination of sample
			HemaSpot HD	Spot On Sciences (San Francisco, United States)	moisture-tight cartridge incorporated with sampling membrane and desiccant	~160 μ L	Multiple punches	opportunity to run a variety of assays from a single sample; patented membrane said to reduce hematocrit effect	potential volume bias
			HemaXis DB10	DBS Systems SA (Gland, Switzerland)	microfluidic chip with a panel of microfluidic channels and standard Whatman filter card encased in a protective plastic cassette	10 μ L	4	compatible with automated lab-processing systems	easy contamination of sample due to open device format; incomplete channel filling or transfer; incompatibility with blood containing organic solvents
Dried Matrix Spot	Dried Blood Spot	Non-Volumetric Dried Blood Spot	Capitainer B	Capitainer AB (Stockholm, Sweden)	microfluidic channels with inlet and outlet ports with dissolvable films	10 μ L	2	eliminates risk of over- or under-filling by integrating successful sampling indicator	possibility of incomplete channel filling
			HemaPEN	Trajan Scientific and Medical (Melbourne, Australia)	K2 EDTA-coated microcapillaries and pre-punched DBS discs housed in a transparent housing and a base	2.74 μ L	4	No sample contamination due to closed housing	requires manual labor and specialized hemaPEN opening tool to retrieve samples

(Continues)

TABLE 1 (Continued)

Technology	Classification	Sub-classification	Device	Company	Design	Volume	Samples	Advantage	Disadvantage
Dried Plasma Spot			HemaSpot HF	Spot On Sciences (San Francisco, United States)	cardidge containing an application surface with small opening, an underlying absorbent fan-shaped collection matrix with 8 blades, and dessicant	10 μ L	8	eight replicates without the need for punching, allowing for repetitive, reproducible, and multiple testing; innovative HemaForm absorbent paper said to promote uniform distribution	possibility of incomplete blade filling
			Telimmune Card	Telimmune (West Lafayette, Indiana, United States)	card format with spreading membrane, separation membrane, and collection discs	3 μ L plasma from 25 μ L whole blood	1 or 2	volumetric plasma collection from non-volumetric whole blood application; plasma collection without centrifugation	
			HemaSpot SE	Spot On Sciences (San Francisco, United States)	cartridge containing spiral shaped membrane and dessicant		Multiple punches	Separation of whole blood into its constituents without centrifugation	
Volumetric Tip Microsampling			Mitra	Trajan Scientific and Medical (Melbourne, Australia)	plastic handler with a proprietary hydrophilic polymer tip	10 μ L, 20 μ L, or 30 μ L	1	volumetric collection independent of hematocrit	possibility of background signal from the polymer tips; contamination onto plastic handler; Only one sample per tip-multiple samples for testing; expensive
			TASSO-M20	Tasso Inc. (Seattle, United States)	device with button, lancet, microfluidic channels connected to a sample pod with volumetric tips	17.5 μ L	4	volumetric collection independent of hematocrit	possibility of background signal from the volumetric tips;




FIGURE 2 Dried matrix spot technology. (A) Dried blood spot at different levels of blood haematocrit. (B) Patterned dried blood spot with four uniform replicates. (C) Telimmune plasma separation card with cross-sectional view. (D) HemaSpot SE device with spiral membrane to facilitate separation of whole blood into constituent parts. (Adapted from References 50, 56, 69 and 70)

can impact quantitative analytical accuracy.¹³ Haematocrit (Hct), that is, the volume proportion of red blood cells (RBCs) in blood is the predominant issue of DBS as it influences DBS spot size, spot homogeneity and extraction recovery of analytes.⁴⁹ Blood with higher Hct levels spreads more slowly on filter paper due to increased blood viscosity, producing a smaller spot than blood with lower Hct levels for the same volume of blood (Figure 2A).⁵⁰ This spot bias results in unreliable volumes in fixed-diameter punches impacting the accuracy of analyte quantitations.^{49,51} Further analytical biases are introduced by the inherent component of DBS technology that enables the drying of blood: the filter paper. Filter paper properties determine the maximum loading capacity, blood spreadability, chromatographic effects, analyte stability and recovery. During the formation of DBS, the content of blood droplets may undergo a chromatographic effect or coffee-ring effect due to differential diffusion across the filter paper. Besides Hct, other factors such as humidity, drying conditions and material of the filter paper also contribute to the uneven distribution of analytes.⁵² A variety of filter papers with different functionalities are commercially available: (i) standard untreated cellulose papers such as Whatman 903 and Ahlstrom 226, (ii) pre-treated cellulose papers such as Whatman FTA Elute, FTA DMPK-A and FTA DMPK-B for enzyme inhibition or protein denaturation,⁵³ (iii) cellulose-based variants such as water-soluble carboxymethyl cellulose paper to improve protein precipitation⁵⁴ and (iv) non-cellulose-based alternatives such as hydrophilic-coated woven polyester fibres to provide spot sizes independent of Hct.⁵⁵

Baillargeon et al. developed a patterned DBS (pDBS) card that uses hydrophobic wax barriers to define distinct areas for sample addition,

distribution and storage, thereby regulating sample application, distribution and volume control.⁵⁶ These pDBS cards provide an enhanced sampling technology to obtain four uniform replicates independently of the Hct over a broad range of clinical values (20–60%; Figure 2B). Compared to standard DBS cards, the use of pDBS demonstrated a threefold improvement in the accuracy of haemoglobin estimation at low Hct (20%). Additionally, pDBS revealed no statistically significant difference between the recovery of sodium and certain amino acids in dried versus liquid blood samples. A potential downside of their pDBS is the incompatibility of the wax barriers with organic solvents (e.g. methanol, acetonitrile) and harsh surfactants (e.g. Triton X-100, sodium dodecyl sulphate) which could interfere with accurate sample analysis if leached into the sample during extraction.⁵⁶ A commercially-available microsampling device that allows for multiple punches from a single sample, similar to the pDBS technology, is HemaSpot™ HD.⁵⁷ HemaSpot™ HD consists of a large membrane which reduces the Hct effect and enables the collection of DBS of ~160 µL. The greater surface area provides an opportunity to run a variety of assays on a single sample via multiple punches. The membrane is incorporated in a moisture-tight cartridge with a built-in desiccant to facilitate a robust storage and shipping solution.⁵⁷

Although different modifications to traditional DBS cards exist to minimise the Hct effects, the easiest approach to eliminate the Hct bias related to spot size and inhomogeneity is to analyse the complete DBS spot formed from a volumetric application of blood. Volumetric DBS can be obtained either by punching the entire DBS after the volumetric application of blood or by volumetrically applying blood on pre-punched discs.⁴⁹ Accurate volumes of blood for application


can be procured using a micropipette or microfluidic channels. However, effective pipetting requires skilled personnel, which limits its scope of application due to the reduced feasibility of self-sampling. In contrast, microfluidic channels are increasingly being used in the development of self-sampling-compliant microsampling devices to produce volumetric DBS for analysis. HemaXis™ DB 10,⁵⁸ Capitainer® B⁵⁹ and HemaPEN®⁶⁰ are commercially-available devices for volumetric DBS collection that operate on the principle of passive volumetric control using microfluidic channel geometry. Blood obtained from a finger prick by lancet is introduced to the microfluidic channel for initiation of DBS collection. The contents of the filled channels are subsequently transferred onto a standard filter card in HemaXis™ DB 10 and pre-punched paper discs in Capitainer® B and HemaPEN®. Another device for volumetric DBS collection is the HemaSpot™ HF which is based on the principle of absorbent paper membrane. The fan-shaped design of the collection matrix with eight blades made up of the HemaForm™ absorbent paper promotes the uniform distribution and volumetric collection of blood droplets applied to its surface.⁶¹

Although these microsampling devices provide a robust approach for collecting volumetric DBS samples to overcome the Hct spot size and inhomogeneity bias, they fail to eliminate the impact of Hct on the extraction efficiency of analytes. DBS with high Hct tends to create a barrier for extraction, resulting in decreased recovery and an under-estimation of analyte concentration.^{49,62} Therefore, optimization of extraction procedures and evaluation of analyte recovery and quantitation accuracy at different Hct levels are absolute necessities during DBS assay validation to ensure that biases are within acceptable limits. Carniel et al. evaluated the influence of Hct at levels 30%, 40% and 50% on the accuracy and recovery of clozapine and norclozapine from DBS samples.²⁹ The extraction efficiency was observed to be in the range of 63–67% for clozapine and 58–69% for norclozapine with no significant impact of Hct. The accuracy over all Hct levels was between 98% and 105%, which was within acceptable limits. Other recent studies have also determined that the effect of Hct on DBS sampling is within acceptable limits for a variety of analytes; 8-isoPGF_{2α}, 8-isoPGE₂ and PGE₂ [Hct 30–60%],²² tacrolimus [Hct 18–55%],^{63,64} creatinine [Hct 18–55%],⁶³ mycophenolic acid [Hct 23–53%],⁶⁴ phenylalanine and tyrosine [Hct 29–64%]²⁵ and vancomycin, meropenem and linezolid [Hct 15–40%].⁶⁵ However, research by Chiu et al. on the influence of Hct ranging from 15% to 65% on the quantification accuracy of four monoclonal antibodies (mAb) demonstrated a need for Hct-based concentration correction to obtain accurate quantification results.³¹ The study exhibited a positive bias exceeding 50% for Hct lower than 25% and a negative bias exceeding 30% for Hct higher than 55%. The bias was attributed to the large size of therapeutic mAbs, which prevented them from passing through the cell membranes of erythrocytes, resulting in their predominant distribution in plasma and an obvious Hct effect. As demonstrated by these different studies, utilization of DBS samples necessitates analyte-specific validation for the Hct effect. Analyte-specific validation for Hct effect is particularly important because Hct levels can vary significantly among individuals and even within individuals at different times.⁶⁶ The validation would ensure that laboratory test results are accurate, even in samples

with varying Hct levels, allowing clinicians to make informed decisions based on reliable diagnostic information to provide high-quality patient care. However, analyte-specific validation requires analysis of samples at various Hct levels, which can be time-consuming and labour-intensive. The process would incur additional testing expenses and require specialised equipment for the accurate measurement of Hct.

Despite the inconveniences caused by Hct, several advancements in the DBS technology have been made to improve existing features and develop new functionalities. A new DBS method was recently developed by Ten-Doménech et al. for accurate, simultaneous quantification of oxidised and reduced glutathione by addressing the challenge of glutathione instability caused by rapid enzymatic or non-enzymatic oxidation in-sample.²³ In this study, a 6-mm-diameter filter paper was pre-treated with 8 µL of N-ethylmaleimide before DBS sampling to allow immediate derivatization of glutathione at room temperature. This on-spot derivatization technique prevented the oxidation of glutathione-to-glutathione disulphide in samples, circumventing the need for manual intervention by clinical staff for sample derivatization. The method was validated for two clinical scenarios: short-term storage using residual blood volumes from newborn screening processed after storage at 4°C for 24 h and long-term storage using cord blood samples processed after storage at –20°C within 1 month. The accuracy and precision of glutathione and glutathione disulphide measurements using this approach were comparable to those of conventional liquid-blood analyses. However, technological advances are application-specific and must be validated to ensure comparable outcomes to conventional approaches. For example, Nguyen et al. compared the performance of a smart DBS sampler for on-paper trypsin digestion to that of the traditional in-solution trypsin digestion and found the latter to be superior for proteomic profiling.⁶⁷ Although the smart sampler, created by immobilizing trypsin on KIO₄-functionalised DBS filter paper, demonstrated that covalently-bound trypsin had significantly better stability with low autolysis, its performance was discovered to be protein-dependent. Digestion of a simple protein, such as cytochrome C, produced similar amounts of peptides in both on-paper digestion and in-solution digestion, whereas digestion of proteins with disulphide bridges, such as bovine serum albumin, produced more peptides in solution. Due to immediate digestion occurring before reduction and alkylation in the smart sampler, on-paper digestion displayed lower protein coverage compared to in-solution digestion.

An alternative to eliminating the Hct effect is the collection of plasma or serum instead of whole blood onto the filter paper. The dried plasma spot (DPS) or dried serum spot (DSS) allows the collection and preservation of the liquid fraction of blood on filter paper without interference from the cellular components.⁶⁸ Similar to DBS, DPS and DSS offer several advantages over traditional plasma or serum samples, including improved stability, ease of collection and transportation, reduced costs and compatibility with automated analysis. However, the conventional DPS and DSS technologies require centrifugation to produce plasma or serum, which can be an expensive and time-consuming component in any workflow. Alternatively, the collection of plasma or serum from whole blood can be enabled



Recent advances in spheroid technology include the development of a spheroid-based microsampling device by Frey et al. to facilitate volumetric blood collection with short drying durations.⁷⁴ In this study, before the silanization process, the paper substrate was embossed with 3D-printed templates to create 20 μL wells to collect whole-blood spheroids. The device included a drying chamber to house the paper substrate containing blood spheroids and 50 μL of ethanol solution for rapid drying facilitated by the increased rate of evaporation of ethanol-water mixtures. Each embossed well incorporating a dried blood spheroid was punched out, inverted onto a second triangular hydrophobic paper and punctured using a sharp 16-gauge hypodermic sterile needle to provide access to the blood components existing in the semi-solid state, for pinhole PSI-MS/MS analysis (Figure 3B). Cocaine and its metabolite benzoylecgonine were used as model chemical systems to evaluate the device because cocaine had been shown to be labile in DBS.⁷¹ The approach was validated by comparing conventional hydrophilic DBS coupled to PSI-MS/MS and traditional LC-ESI-MS/MS with liquid-liquid extraction. When compared to samples made with planar hydrophilic paper, the sample stability and sensitivity ($\sim 10\times$) were found to be superior in dried blood spheroids.⁷⁴ Additionally, LC-ESI-MS/MS experiments were found to be $\sim 10\times$ more sensitive with lower LOD and LOQ (cocaine: 0.015 and 1.33 ng/mL; benzoylecgonine: 0.028 and 2.22 ng/mL) compared to pinhole PSI-MS/MS experiments (cocaine: 0.12 and 2.88 ng/mL; benzoylecgonine: 0.49 and 4.43 ng/mL). The lower sensitivities are primarily attributable to the intrinsic matrix effects resulting from the direct analysis of whole blood in PSI-MS/MS as compared to the reduced matrix effects resulting from prior sample clean-up procedures in LC-ESI-MS/MS. The clinical efficacy of the spheroid technology was further demonstrated by Sham et al.




FIGURE 4 Volumetric tip microsampling technology. (A) Mitra device by Neoteryx. (B) TASSO-M20 device by TASSO Inc. (Adapted from References 75 and 76)

who measured creatinine levels in whole-blood spheroid microsamples obtained from adult human volunteers.²¹ The study compared the PSI-MS/MS analysis of 5 μ L of oven-dried whole blood spheroid collected on a planar hydrophobic paper to the conventional UPLC-MS/MS analysis of 50 μ L of liquid whole blood. Similar results were obtained from both approaches with a maximum deviation of 0.3 μ g/mL. The spheroid approach exhibited excellent linearity ($R^2 > 0.99$; 2.5–20 μ g/mL) across the expected human concentration range, a lower limit of quantification of 2.5 μ g/mL, precision $\leq 6.3\%$ and recovery in the range of 88–94%.²¹

2.1.3 | Volumetric tip microsampling

Volumetric tip microsampling (VTM), introduced to the market in 2014, is an alternative microsampling technique to collect Hct-independent, fixed-volume dried whole-blood samples by accurate and precise absorption of blood into a volumetric polymer tip. The volume of blood collected is controlled by the amount and properties of the polymeric substrate. After the samples collected on these tips are dried, they are stored or transported to the laboratory for analysis.¹⁴ The VTM technique is currently employed by two commercially-available blood collection devices: Mitra and TASSO-M20. The Mitra device comprises a plastic sampler attached to a proprietary hydrophilic polymer tip which functions based on their patented volumetric absorptive microsampling (VAMS®) technology. To initiate sample collection, the leading surface of the Mitra tip is dipped into blood obtained via a finger prick (Figure 4A).⁷⁵ The Mitra devices are compatible with 96-Autoracks and automated liquid handling systems which facilitate high-throughput testing. The TASSO device consists of a large button, a lancet, microfluid channels and a sample-collection pod with four TASSO volumetric tips. For TASSO-M20 blood collection, the device adheres to the upper arm of the patient and the button is pressed firmly and released, which creates a vacuum and deploys the lancet

to puncture the arm to draw blood which is directed to the tips via the capillary channels (Figure 4B).⁷⁶ With the use of VTM for sample collection, issues of inhomogeneity experienced by DBS sampling are eliminated, as there is no filter paper component to cause the biases, such as uneven spreading, distribution and sample size variability.

However, certain aspects of the VTM technology require careful consideration prior to its implementation. A critical and comprehensive overview of the practical aspects of VAMS can be found in the 2018 review by Kok et al.⁷⁷ One of the potential challenges of VTM is the difficulty in the extraction of analytes which could be trapped in the crevices of the porous polymer tip due to the possibility of dried erythrocyte occlusion.²⁵ However, studies have shown that additional processes such as pre-wetting the polymer tip with water and sonication during sample preparation can improve the extraction efficiency of analytes.^{77,78} Another limitation with VTM is the inability to add stabilizing agents during sample collection for the accurate quantification of unstable analytes susceptible to hydrolytic degradation. Addition of reagents to the polymer tips prior to sample collection could potentially result in undesirable sample volumes and inhomogeneous distribution of analytes. Nevertheless, different drying conditions, such as the use of a desiccator, pressurization, nitrogen gas and household vacuum sealer, have been shown to improve sample stability. In the analysis of acetylsalicylic acid, a hydrolysis-prone analyte, Moon et al. showed that humidity control was critical for its stability, and the use of a simple household vacuum sealer could maintain the stability for up to 1 month.³³

Similar to DBS filter paper modifications, the polymeric tips can be modified to attain new features. Comparable to the smart DBS sampler for proteomics, Reubsat et al. modified the Mitra device to build a smart VAMS proteolysis sampler. Maximum proteolytic sampler activity was achieved when at least 50 μ g of trypsin was covalently bound to the Mitra tip at a pH of 9. The study observed no significant difference in the absorbed sample volume or drying rate even though the tips were subjected to chemical modifications before sampling. A

TABLE 2 Overview of commercially-available liquid matrix microsampling devices.

Technology	Device	Company	Design	Volume	Samples	Advantage
Capillary-based	Minivette POCT	Sarstedt AG & Co. KG (Numbrecht, Germany)	Plastic volumetric capillaries	10–200 μ L	1	Collection of volumetric liquid whole blood; compatible for volumetric DBS collection
	Microvette APT	Sarstedt AG & Co. KG (Numbrecht, Germany)	Plastic capillary or a collection rim	250–500 μ L	Multiple	Tubes compatible with automated sample processing in blood count analysis systems
	Microsampling Wing	Shimadzu Europa (Duisburg, Germany)	Olefin polymer U-shaped capillary microchannel with H-shaped segment at the top with an inlet port	23 μ L whole blood (5.6 μ L plasma (2.8 μ L + 2.8 μ L))	1 or 2	Easy plasma fractionation; low human error
Tube-based	TASSO+	Tasso Inc. (Seattle, United States)	Device with button, lancet, microfluidic channels connected to standard collection tubes	250–600 μ L	Multiple	Compatible with multiple standard collection tubes
	TASSO-SST	Tasso Inc. (Seattle, United States)	Device with button, lancet, microfluidic channels connected to tube with serum separator gel with no anticoagulants	200–300 μ L	Multiple	Collection of serum
	TAP II	YourBio Health, Inc. (Medford, United States)	Device with button, solid microneedle array, connected to standard collection tubes	300–900 μ L	Multiple	Compatible with multiple standard collection tubes; skin puncture via solid microneedles-less painful

comparison of in-solution trypsin digestion with the VAMS digestion for seven proteins revealed higher intensities of tryptic peptides in VAMS digestion for four proteins. The smart sampler was also applied to complex serum samples with overnight digestion, which yielded similar peptide and protein identifications to in-solution digestion, demonstrating the potential of modified VAMS for proteomics sample preparation.⁷⁹

2.2 | Liquid matrix microsampling

An overview of commercially-available liquid matrix microsampling devices can be found in Table 2.

2.2.1 | Capillary microsampling

Capillary microsampling (CMS) is a versatile and easily implemented technique for the collection of small, accurate volumes of liquid matrices, including whole blood.⁸⁰ The CMS technology enables the collection of accurate blood volume using exact-volume capillaries,

usually coated with anticoagulants, to take up blood obtained via finger prick through capillary forces. The filled capillaries can be stored in sample tubes for later use or prepared for immediate use. One of the commonly employed pre-treatment procedures is centrifugation for the generation of plasma. Typically, the capillaries are sealed with wax to prevent sample loss during centrifugation. After centrifugation, the plasma samples can be transferred using a micropipette into a sample tube or an end-to-end transfer into another capillary.⁸¹ CMS is compatible with well-established sample-preparation techniques, including the addition of agents for immediate stabilization of unstable compounds, and laboratory automation.⁸⁰ Some of the commercially-available CMS devices are the Sarstedt Tubes, such as the Minivette® POCT series and Microvette® APT, and the Microsampling Wing (MSW). The Minivette® POCT series consists of plastic capillaries which are available in a range of sample volumes (10 μ L–200 μ L) and anticoagulants for preparation to facilitate the simple collection and dispensing of blood for point-of-care tests (Figure 5A).⁸² The Microvette® APT enables the collection of K₂ EDTA whole blood (250–500 μ L) directly into the tube for routine analysis either via end-to-end capillary filling or a collection rim. It is specially designed to facilitate automated sample processing in blood count analysis




FIGURE 5 Liquid matrix microsampling technology. (A) Minivette® POCT. (B) Microvette® APT. (C) Microsampling wing and microsampling windmill. (D) TASSO+. (E) TASSO-SST. (F) TAP II. (Adapted from References 82–87)

systems (Figure 5B).⁸³ The MSW device comprises a U-shaped capillary microchannel with an H-shaped segment at the top containing EDTA anticoagulant for the volumetric collection of whole-blood samples which can be introduced through an inlet port. Upon centrifugation of the device, the plasma is retained in the H-shaped segment which can be snapped into two separate segments to provide two volumetrically accurate plasma samples. Additionally, a specialised holder called the Microsampling Windmill allows for the simultaneous and efficient centrifugation of 14 MSWs at a time (Figure 5C).⁸⁴

2.2.2 | Tube microsampling

Tube microsampling (TMS) involves the collection of liquid blood microsamples directly into microtubes. Three commercially-available devices use the TMS technology: TASSO+, TASSO-SST and TAP II devices (Figure 5D–F). The design and operation of these TMS-based devices are similar to the TASSO-M20 device. Instead of the sample pod, the TASSO+ and TASSO-SST devices consist of appropriate microtubes with or without anticoagulants to facilitate the downstream collection of plasma (TASSO+) or serum (TASSO-SST).^{85,86} These devices are compatible with a variety of additive microtubes. The difference between the TAP II device and the TMS-based TASSO devices is the skin-puncturing mechanism during the blood collection procedure. Instead of a button-activated lancet deployment, the TAP II device deploys a solid microneedle array for skin puncture and a vacuum for blood flow into the microtubes.⁸⁷ Microneedles are known to be less painful than lancets for the skin puncture due to

their micron-sized dimensions which prevent them from hitting the nerve endings when inserted into the skin.⁸⁸ These single-use, button-activated devices enable at-home self-sampling of up to 500 μ L of blood into microtubes which can be shipped to the laboratory for further analysis. Larger volumes of blood may be collected by these devices in the future if their designs are modified to create stronger vacuum or larger negative pressure.

3 | APPLICATIONS

The following section summarises microsampling research published in the year 2022 in the field of volume-limited applications, TDM, biomarker research, forensic toxicology and sports anti-doping. An overview of the research articles on method development and validation studies can be found in Table 3.

3.1 | Volume-limited applications

3.1.1 | Animal studies

Small animals are widely employed in pre-clinical research for drug development for PK, PD, toxicokinetic (TK) and toxicodynamic (TD) investigations to determine the bioavailability of drugs at pharmacological concentrations and ensure drug safety. These studies typically require collection of blood samples from a single animal candidate at several timepoints, which can lead to health deterioration or even

TABLE 3 Overview of microsampling research in biomedical applications published in the year 2022.

Application	Technology	Device	Sample volume	Target compound	Analytical method	Reference
Animal studies	DBS using CMS	Glass capillary and Whatman 903	~	Clozapine	LC-MS	[94]
	VTM	Mitra	10 µL	Vincristine and tariquidar	LC-MS/MS	[27]
	CMS	Heparinised capillary tubes	10 µL	E7130	UHPLC-HRMS	[93]
	CMS	Glass microcapillaries	15 µL	Trans-resveratrol and its main metabolites	HPLC-MS/MS	[92]
Neonatal and paediatric applications	~	~	10 µL	Thiotepa, tepa, cyclophosphamide (CP), and 4-hydroxycyclophosphamide (4-OHCP)	LC-MS/MS	[102]
	DBS	Whatman 903	8 mm punch (~17 µL)	Gentamicin sulfate	UHPLC-MS/MS	[100]
	DBS	Whatman 903, Minivette POCT, HemaXis, Capitainer	6 mm punch, 20 and 50 µL, 10 µL, 13.5 µL	Isoprostanooids 8-iso prostaglandin F2α and 8-iso prostaglandin E2, and prostaglandin E2	UHPLC-MS/MS	[22]
	DBS	~	10 µL	Oxidised and reduced glutathione	LC-MS	[23]
	DBS	Perkin Elmer 226, HemaXis, Capitainer, Mitra	3.2 mm punch and varying volumes tested	Phenylalanine and tyrosine	FIA-MS/MS	[25]
	VTM	Mitra	10 µL	Gentamicin	LC-MS/MS	[101]
	VTM	Mitra	10 µL	Furosemide	UHPLC-MS/MS	[103]
	CMS	Microsampling Wing and Li-H capillary tube	5.6 µL	Meropenem	LC-MS/MS	[30]
	DBS and CMS	Capillary and Whatman 903	50 µL and 10 mm punch	Vancomycin, meropenem and linezolid	UHPLC-MS/MS	[65]
	CMS	EDTA coated capillary	70 µL	Tafenoquine	LC-MS/MS	[105]
Therapeutic drug monitoring	CMS	heparinised plastic capillary tubes	50 µL	Cefotaxime and desacetylcefotaxime	UHPLC-MS/MS	[26]
	VTM	TASSO-M20	20 µL	Tacrolimus and mycophenolic acid	LC-MS/MS	[113]
	DBS and VTM	Whatman 903 and Mitra	3 mm punch and 20 µL	Tacrolimus and creatinine	LC-MS/MS	[63]
	DBS and VTM	HemaXis and Mitra	10 µL	Clozapine	HPLC-ED	[126]
	VTM	Mitra	~	Hydrochlorothiazide, furosemide, lisinopril, and torsemide, as well as the metabolites canrenone, enalaprilat, and ramiprilat	LC-MS	[127]
	VTM	Mitra	20 µL	Methotrexate and methotrexate polyglutamates [MTX-PG ₁₋₅ and MTX-PG ₆₋₇]	LC-MS/MS	[28]
						(Continues)

TABLE 3 (Continued)

Application	Technology	Device	Sample volume	Target compound	Analytical method	Reference
DBS and DPS		HemaXis and Whatman 903	10 µL and 6 mm punch	clozapine and norclozapine	HPLC-MS/MS	[29]
VTM		Mitra	30 µL	Cannabidiol (CBD), tetrahydrocannabinol (THC) and their respective metabolites: cannabidiol-7-oic acid (7-COOH-CBD); 7-hydroxy-cannabidiol (7-OH-CBD); 6- α -hydroxy-cannabidiol (6- α -OH-CBD); and 6- β -hydroxycannabidiol (6- β -OH-CBD); 11-Hydroxy- Δ^9 -tetrahydrocannabinol (11-OH-THC) and 11Nor-9-carboxy- Δ^9 -tetrahydrocannabinol (THCCOOH)	UHPLC-MS/MS	[131]
DPS		~	~	methotrexate	LC-MS/MS	[117]
DBS and VTM		Whatman 903 and DMPK-C, HemaXis, Capitainer-B, and Mitra	~	Tacrolimus, ciclosporin, everolimus, sirolimus and mycophenolic acid	LC-MS/MS	[115]
DBS		Whatman 903	15 µL	Bevacizumab, trastuzumab, nivolumab and tocilizumab	LC-MS/MS	[31]
DBS using CMS		Capillary micropipette and Whatman 903	10 µL	Tacrolimus, sirolimus, everolimus and cyclosporin A	LC-MS/MS	[112]
TMS		TASSO-SST	~	CRP, RF IgM, CCP IgG	Phadia™ 250 instrument and Thermo Scientific™ Indiko™ Plus Clinical Chemistry Analyzer	[24]
DBS and VTM		Whatman 903 and Mitra	6 mm punch and 20 µL	Tacrolimus and mycophenolic acid	LC-MS/MS	[64]
VTM		Mitra	20 µL	Imatinib mesylate	LC-MS/MS	[125]
VTM		Mitra	10 µL	Capecitabine, 5'-deoxy-5-fluorocytidine, 5'-deoxy-5-fluorouridine and 5-fluorouracil	LC-MS/MS	[40]
VTM		Mitra	10 µL	Bosutinib, dasatinib, gilteritinib, ibrutinib, imatinib, midostaurin, nilotinib and ponatinib	LC-MS/MS	[78]
VTM		Mitra	20 µL	Axitinib	LC-MS/MS	[120]

(Continues)

TABLE 3 (Continued)

Application	Technology	Device	Sample volume	Target compound	Analytical method	Reference
	VTM	Mitra	20 µL	Cabozantinib, dabrafenib, nilotinib, osimertinib, afatinib, axitinib, bosutinib, lenvatinib, ruxolitinib and trametinib	LC-MS/MS	[121]
	VTM	Mitra	~	Acetylsalicylic acid	LC-MS/MS	[33]
	DBS and VTM	HemaXis and Mitra	10 µL and 20 µL	Tacrolimus, mycophenolic acid, creatinine and iohexol	LC-MS/MS	[32]
	VTM	Mitra	10 µL	Paracetamol and its four metabolites: paracetamol-glucuronide (APAP-Gluc), paracetamol-sulphate (APAP-Sulf), paracetamol-mercapturate (APAP-Merc), and paracetamol-cysteine (APAP-Cys)	LC-MS/MS	[128]
	DPS	Telimmune Card	~70 µL	Giredestrant	LC-MS/MS	[118]
	VTM	Mitra	~	Favipiravir	HPLC-Diode Array Detector	[129]
	VTM	Mitra	30 µL	Cyclophosphamide (CP), and 4-hydroxycyclophosphamide (4-OHCP)	UPLC-MS/MS	[119]
	VTM	Mitra	20 µL	Mercury	DMA80-evo® Analyzer	[41]
	DBS	Whatman 903	30 µL and 50 µL	Mercury	Thermal Decomposition Amalgamation and Atomic Absorption Spectrometry (TDA-AAS) in a Direct Mercury 18 Analyzer (DMA)	[145]
	VTM	Mitra	30 µL	Antibodies against SARS-CoV-2 nucleocapsid (IgG), spike protein (IgG, IgM, IgA), and receptor binding domain (IgG, IgM, IgA)	enzyme linked immunosorbent assay	[144]
Omics-based biomarker research	DBS and VTM	HemaXis and Mitra	10 µL	SARSCoV-2 anti-Spike IgG antibody	microfluidic nano-immunoassay	[143]
	TMS	TASSO-SST	~	17 protein inflammatory biomarkers including CRP, Ferritin, IL-6, PCT, D-dimer, IL-1B and IL-1Ra	Ella multi-analyte immunoassay	[42]
	DBS and VTM	Whatman 903 and Mitra	20 µL	aflatoxin B1	LC-MS/MS	[141]
						(Continues)

TABLE 3 (Continued)

Application	Technology	Device	Sample volume	Target compound	Analytical method	Reference
	VTM	Mitra	10 μ L	proteomic profile	LC-MS	[134]
	DBS and DPS	Whatman 903 and Telimmune card	6 mm punch and 3.2 μ L	498 compounds covering 24 lipid subclasses	UHPLC-MS/MS	[132]
	VTM	Mitra	~	1600 proteins	LC-MS	[135]
	VTM	Mitra	10 μ L	24 compounds related to tryptophan metabolism	LC-MS/MS	[138]
	TMS	TASSO-SST	~	Uracil	LC-MS/MS	[137]
	VTM	Mitra	30 μ L	Acrylamide and glycidamide	HPLC-UV	[140]
	VTM	Mitra	~	Testosterone, cortisol, 25 hydroxyvitamin D and bone resorption marker β -CTX	LC-MS/MS and COBAS 6000 e601 using electro- chemiluminescence immunoassay	[139]
	VTM	Mitra	10 μ L	35 metabolites including taurine, serotonin, hypoxanthine, glutathione, uric acid, 1-methyl histidine, acetyl-carnitine, and hypoxanthine	LC-MS	[136]
	Dried Blood Spheroid	~	5 μ L	Creatinine	PSI-MS/MS	[21]
	DBS and VTM	Whatman 903 and Mitra	20 μ L	Continuous erythropoietin receptor activator (CERA)	ELISA kit	[38]
Forensic toxicology	Dried Blood Spheroid	~	20 μ L	Cocaine and its metabolite benzoylecgonine	PSI-MS/MS	[74]
	DBS	Whatman FTA TM	20 μ L	425 drugs including including benzodiazepines, antipsychotics, antidepressants, antipyretic analgesics, non-steroidal anti-inflammatory drugs, antibiotics, antiepileptic drugs, and new psychoactive drugs	LC-MS/MS	[153]
	DBS and VTM	Whatman 903 and TASSO-M20	60 μ L	Endogenous erythropoietin (hEPO) and prohibited ERAs (BRP, NESP, CERA and EPO-Fc)	SAR-PAGE and Western blotting	[37]
	DBS	~	30 μ L	PEth	UHPLC-QTOF- HRMS and UHPLC-MS/MS	[148]
	DBS and VTM	HemaXis and Mitra	~	Peth 16:0/18:1, PEth16:0/20:4 and ethylglucuronide (EtG)	LC-MS/MS	[150]
	DBS	~	~	13 anabolic steroid esters	DBSA-TLX-HRMS	[158]
						(Continues)

TABLE 3 (Continued)

Application	Technology	Device	Sample volume	Target compound	Analytical method	Reference
	DBS using CMS	Whatman 903 and capillary	20 µL	PEth-homologues (16:0/18:1; 16:0/18:2; 16:0/20:4; 17:0/18:1; 18:0/18:1; 18:1/18:1; 18:0/18:2)	LC-MS/MS	[151]
	DBS	HemaXis	10 µL	PEth 16:0/18:1	LC-HRMS	[149]
	DBS	Whatman	30 µL	ricinine and L-abrine	LC-MS/MS	[155]
	DBS and TMS	~	~	Aluminium (Al), total arsenic (As), cadmium (Cd), cobalt (Co), chromium (Cr), copper (Cu), manganese (Mn), molybdenum (Mo), nickel (Ni), lead (Pb), selenium (Se) and zinc (Zn)	ICP-MS	[156]
	DBS	FTA™ DMPK C	30 µL	132 synthetic opioids, cathinones, hallucinogens & fentanyl	UHPLC-HRMS-QTOF	[34]
	DBS and VTM	HemaSpot-HF, Whatman 903 Protein Saver cards, Whatman FTA DMPK-A, B and C cards, Tasso-M20, and Mitra	Varying volumes	235 drugs	UHPLC – HRMS	[39]
	DBS	HemaXis	10 µL	PEth 16:0/18:1	LC-MS/MS	[35]
	DBS	HemaPEN	2.74 µL	Cocaine and the metabolites benzoylecgonine and cocaethylene	LC-MS/MS	[152]
	VTM	Mitra	20 µL	85 licit and illicit drugs	LC-HRMS	[154]

CCP, cyclic citrullinated peptide; CMS, capillary microsampling; CRP, C-reactive protein; DBS, dried blood spot; DPS, dried plasma spot; ED, electrochemical detection; FIA, flow injection analysis; HRMS, high-resolution MS; ICP, inductively coupled plasma; Ig, immunoglobulin; IL-1 β , interleukin-1 beta; IL-Ra, interleukin-1 receptor antagonist; Li-H, lithium-heparin; PCT, procaltitonin; PSI, paper spray ionisation; RF, rheumatoid factor; UHPLC, ultra-high performance LC; VTM, volumetric tip microsampling; '~', not available.

early death, due to extensive loss of blood. To protect the well-being of animals used in research, the Universities Federation for Animal Research advocated the 3Rs principle: Replace, Reduce and Refine.⁸⁹ Making animal studies more ethical through the 3Rs can be achieved with microsampling, as it reduces stress, sampling time and the number of animals needed for studies.⁹⁰ Guidelines that consolidate the regulatory perspectives on microsampling and support its use in blood sampling have also been added to and reviewed by the International Council for Harmonization of Technical Requirements for Pharmaceuticals for Human Use.⁹¹

Different microsampling technologies have been employed in recent investigations by Rosser et al., Xu et al., Ryu et al. and Pelcovà et al. in animal studies for PK applications. In the work of Rosser et al., the PK assessment of vincristine and tariquidar in mice using an LC-MS/MS approach was effectively carried out using the VAMS technology. Assessment of both single-agent therapy and combination therapy over a 24 h period in 10 µL microsamples revealed a 2.3-fold increase in vincristine drug exposure when combined with tariquidar, which validated the use of the approach for longitudinal analysis of drug exposure in animal studies.²⁷ Xu et al. evaluated the feasibility of CMS coupled to HPLC-MS/MS for mice PK studies using trans-resveratrol as the model drug. The PK parameters of the drug and its associated metabolites found in the study were comparable with those reported in previous studies, which demonstrated that CMS provides credible samples for subsequent quantitative PK analysis.⁹² Similarly, Ryu et al. also successfully employed CMS in the PK study of E7130 in mice using an HRMS quantification method.⁹³ Additionally, a local and national animal welfare committee approved PK pilot study of clozapine was successfully conducted on a single albino rat using DBS samples spotted using CMS.⁹⁴

Traditionally, satellite animals are used in TK analysis to avoid the effects of blood sampling on toxicological endpoints in main study animals, which leads to a higher number of required study animals. Microsampling has been shown to facilitate toxicity assessment and TK analysis in the same animals, even for hematotoxic compounds. Implementation of microsampling in TK studies is often contested by regulatory bodies due to the unidentified effects of microsampling on TK parameters, necessitating bridging studies.⁹⁵ Tochitani et al. conducted a 2-week rat study to examine the effects of microsampling on toxicity assessment using methylene blue trihydrate and azathioprine as test compounds. Using urinalysis, haematology, blood chemistry, organ weights and histopathology, the study compared the results of toxicological end points between non-microsampling and microsampling subgroups. The differences between the subgroups were found to be small and insignificant, indicating that toxicity of hematotoxic compounds could be detected even with microsampling without any overestimation of toxicity.⁹⁶ In another study, during the examination of 109 carcinogenicity studies undertaken for new drug applications involving a minimum of 65,341 mice and rats, Manuppello et al. determined that an 18.7% reduction in animal size could have been accomplished by evaluating exposure by microsampling blood in main study animals as opposed to TK satellites.⁹⁷

3.1.2 | Neonatal and paediatric studies

Research in neonatal and paediatric populations is challenging due to their small volume of circulating blood and the physiological and psychological burden associated with blood sampling. Pain and stress in the early phase of life are known to affect the neurodevelopment leading to poor cognitive outcomes and long-term effects on social-emotional behaviour.⁹⁸ Blood sampling with automatic lancets has been demonstrated to cause less pain in neonates than with needles.⁹⁹ Furthermore, the physiological and pharmacological characteristics of these populations, in particular the preterm infants, differ significantly from adults, which often necessitate tailored dosing regimens for medication administration. However, standard blood sampling for PK studies and TDM may increase the risk of iatrogenic anaemia in critically ill neonatal and paediatric populations. Microsampling may overcome both the ethical and physiological concerns, improving the feasibility of PK studies in these populations to supersede the use of off-label drugs and dosing regimens extrapolated from clinical studies conducted on adults.

Several assays for the quantification of drugs in microsamples have been developed for use in neonatal and paediatric PK studies. A ultra-high-performance LC (UHPLC) MS/MS method capable of identifying distinct gentamicin forms was recently developed by dos Santos et al., allowing for the measurement of their concentrations ranging from 0.1–40 mg/L in non-volumetric DBS. The stability of gentamicin was determined to be 21 days at –20 and 8°C and 24 h at room temperature. The assay was evaluated with nine neonatal patients which revealed that the difference between the estimated plasma concentration from DBS, determined using the Hct level and a correction factor, and concentration measured using conventional venous plasma was statistically insignificant but highly variable.¹⁰⁰ Similarly, Parker et al. reported for the first time the use of VAMS in neonatal TDM of gentamicin. The study including two neonates demonstrated that the comparison of calculated gentamicin concentration from VAMS to observed plasma concentration met the acceptance criteria of an incurred sample reanalysis test ($\leq 20\%$) and results were obtained promptly.¹⁰¹ Other LC-MS/MS assays that have been recently developed include assays for the quantification of thiotepa, its metabolite tepa, CP and its metabolite 4OHCP in 10 µL liquid plasma samples and furosemide in 10 µL VAMS samples.^{102,103}

As a part of clinical bridging studies, several studies have compared two or more microsampling approaches for a given neonatal or paediatric application. For the analysis of meropenem, a capillary with lithium heparin anticoagulant was found to perform better than the MSW, although both yielded acceptable concentrations comparable to those obtained from conventional arterial catheter sampling.³⁰ In another study, Xiaoyong et al. compared DBS and CMS for the measurement of vancomycin, meropenem and linezolid. The results showed a significant correlation between DBS and CMS measurements ($R > 0.98$; $p < 0.01$) for all three antibiotics. However, CMS to DBS drug concentration ratios were highly variable for vancomycin (~1.39) and meropenem (~1.34) which required an Hct correction

factor for accurate measurements. The study suggests DBS for unstable, hydrophobic compounds, such as meropenem and linezolid, but CMS for vancomycin due to its low adsorption capability and high stability.⁶⁵ Biagini et al. evaluated non-volumetric DBS and volumetric DBS (HemaXis™ DB, Minivette POCT, Capitainer® B) for the determination of isoprostanoids and prostanoids in preterm newborns suffering from Patent Ductus Arteriosus. It was found that the Hct biases affected the non-volumetric DBS measurements. All analytes measured with Minivette POCT exhibited a recovery of 78–108% and a precision of less than 15%, compared to a recovery and precision of 23–81% and 20–25%, respectively, for other volumetric DBS. Based on the analytical performance parameters and cost, the Minivette POCT was found to be the best option for DBS sampling for the application.²² Similarly, volumetric blood collection devices (Capitainer® B, HemaXis™ DB 10 and Mitra) were also evaluated for the monitoring of phenylalanine and tyrosine in patients with phenylketonuria in a study by Carling et al. which revealed their superior analytical performance (mean biases in relation to conventional liquid samples: –7.8 (Capitainer® B), –5.1 (HemaXis™ DB 10), –12 (Mitra)) compared to traditional DBS (–32.6).²⁵

Microsampling can also facilitate data-rich modelling and simulations for population PK research.¹⁰⁴ In a study by Bachhav et al., the population PK models obtained from CMS and venous samples were found to provide the same dosing regimen recommendations for tafenoquine antimalarial in paediatric patients.¹⁰⁵ Guerra Valero et al. also evaluated the suitability of CMS to describe the population PK of cefotaxime and its metabolite desacetylcefotaxime in critically ill paediatric patients.²⁶ The authors further showed that parents and bedside nurses preferred microsampling over conventional sampling for paediatric PK clinical investigations using a seven-point Likert-scale questionnaire.¹⁰⁶ In the field of population-based research, the use of archival newborn screening DBS samples has shown immense potential for emerging targeted and untargeted multi-omics analysis.¹⁰⁷ Microsampling with point-of-care testing has also been identified as a blood conservation strategy that significantly reduces the need for adult RBC transfusion therapy for anaemia of prematurity.¹⁰⁸

3.2 | Therapeutic drug monitoring

TDM specialises in the measurement of drug concentrations in biological fluids, such as serum, plasma or blood, to facilitate personalised medicine. Individualised drug therapy for improved patient outcomes is ensured through the monitoring of drug concentration profiles in patients to adapt drug dosage, interval and duration to account for potential inter-individual genetic and metabolic differences in drug responsiveness and the drug concentration-effect correlation.¹⁰⁹ Thus, TDM implementation requires timely, rapid and accurate measurements of target drug concentrations to evaluate the clinical therapeutic efficacy and prevent failure (too low concentration) and toxicity (too high concentration). Microsampling has been increasingly employed in TDM to facilitate at-home, convenient, pain-free sample collection with easy storage and transport.¹¹⁰ Despite

being a relatively robust method for routine TDM, implementation of microsampling requires development and validation of bioanalytical assays to obtain good analyte sensitivity, stability, extraction efficiency and correct clinical interpretation of results. Routine TDM is prescribed for a variety of therapeutic drug categories such as immunosuppressants (IMS), antibiotics, anticancer drugs, psychoactive drugs, antidepressants and analgesics.¹¹¹

Microsampling has been extensively studied for the determination of tacrolimus and mycophenolic acid, two of the most commonly prescribed IMS used to prevent organ rejection in recipients of solid organ transplants.^{32,63,64,112,113} Studies conducted by Paniagua-González et al. and Zwart et al. in liver or renal transplant patients, respectively, found that although DBS and VAMS, after a concentration correction, provided comparable results to conventional whole-blood tacrolimus and plasma mycophenolic acid measurements, DBS performed better than VAMS in terms of clinical performance (98.7% of paired samples within $\pm 20\%$ difference as opposed to 71.1% of VAMS paired samples), sample quality (1 sample discarded as opposed to 9 VAMS samples due to poor sample quality) and costs (7 times lower than Mitra device).^{32,64} Additionally, use of non-contact Hct prediction strategies such as NIR and UV/VIS spectroscopy coupled to an automated DBS workflow was found to adequately correct for Hct effect in determining IMS concentrations.¹¹⁴ In contrast, for the simultaneous monitoring of tacrolimus and kidney functioning using creatinine, VAMS was found to be the preferred single sampling option compared to DBS, as only 69.4% of paired samples with DBS were within $\pm 15\%$ as opposed to 81.6% with VAMS.⁶³ Additionally, Leino et al. also validated the application of TASSO-M20 for the quantitative bioanalysis of tacrolimus and mycophenolic acid.¹¹³ Although these studies demonstrated the feasibility of microsampling in TDM of IMS, results from a proficiency testing pilot for IMS microsampling assays when compared to whole blood methods showed great interlaboratory variation. Since the results influence clinical decision-making, harmonization and standardization are thus necessary before microsampling techniques can be implemented in patient care.¹¹⁵ A recent review by Deprez et al. summarises the state-of-the-art for DBS-based TDM of IMS, including a critical evaluation of challenges, implementation status in clinical practice and considerations to overcome implementation barriers.¹¹⁶

One of the clinically relevant characteristics of drugs is their differential distribution between cells (erythrocytes and leukocytes) and plasma, which necessitates the selection of an appropriate matrix for TDM analysis. In the treatment of rheumatoid arthritis, the site of action of methotrexate is assumed to be in the peripheral blood mononuclear cells. A highly sensitive and specific LC-MS/MS method was developed for the quantification of methotrexate polyglutamates which are active metabolites of methotrexate. Evaluation of methotrexate polyglutamates in different matrices including plasma, RBCs, peripheral blood mononuclear cells and venous whole blood revealed that VAMS can be used for accurate TDM analysis only if sampling is done immediately before the next dose of methotrexate to ensure that concentrations are not overestimated.²⁸ In another study in patients with lymphoblastic leukaemia, DPS and wet plasma concentrations of methotrexate were found to have no statistically significant

difference.¹¹⁷ However, DPS from Telimmune cards have been shown to produce inconsistent filtration across different cards, in the study by Tang et al., for the quantification of giredestrant, stipulating the need for further optimization to ensure good accuracy and precision.¹¹⁸ Various LC-MS/MS assays for the VAMS technology have been developed and validated for the quantification of other chemotherapy drugs and its metabolites such as 5-fluorouracil, capecitabine, CP, 4-OHCP and several tyrosine kinase inhibitors (TKIs).^{40,78,119–121}

Most available TKIs are susceptible to drug–drug interactions, and their bioavailability is dependent on drug formulation, gastrointestinal absorption and concomitant food intake;^{122,123} therefore, the implementation of TDM of TKIs in the standard care of oncology patients has the potential to improve treatment outcomes and reduce toxicity by monitoring medication adherence and actual drug concentration in plasma. A recent review by Verougstraete et al. provides a comprehensive description of the different methodological aspects to consider in dried blood microsample methods for TKI TDM.¹²⁴ The same authors developed an LC-MS/MS assay for the simultaneous quantification of eight TKIs following substantial optimization to obtain an Hct- and storage- independent sample preparation procedure.⁷⁸ Opitz et al. utilised an on-line SPE system to ease the extraction of axitinib from VAMS samples.¹²⁰ Due to a high blood-to-plasma ratio of TKIs, conversion factors are necessary to compare VAMS TKI concentrations with conventional plasma concentrations. Zimmerman et al. determined 10 TKIs, where nine TKIs exhibited a constant conversion factor over their entire calibration range with only trametinib exhibiting a decreasing conversion factor with increasing VAMS concentration.¹²¹ Further, the degree of patient satisfaction concerning self-collection and treatment adherence using VAMS has been evaluated on chronic myeloid leukaemia patients receiving imatinib mesylate therapy.¹²⁵ The study revealed that 87% of patients experienced less discomfort with VAMS sampling than with conventional sampling but only 77% of patient-collected samples met the quality criteria.

In their respective studies, Carniel et al. and Marasca et al. examined DBS, DPS and VAMS technologies for the determination of antipsychotic drugs, clozapine and norclozapine. DBS and DPS measurements were found to be highly correlated to serum levels, with greater accuracy for DPS compared to DBS.²⁹ A 30-day stability test showed higher stabilities in the dried matrices stored at room temperature when compared to liquid plasma samples stored under the controlled temperature of -20°C .¹²⁶ Further, the VAMS-based strategy was shown to be suitable for assessing adherence to several classes of antihypertensive drugs and quantification of paracetamol and favipiravir.^{127–129} Microsampling has also been applied for the determination of tetrahydrocannabinol, cannabidiol and their respective metabolites in light cannabis smokers and patients with intractable epilepsy.^{130,131} Additionally, an assay was developed to simultaneously quantify four mAbs and estimate Hct using haemoglobin peptides in DBS.³¹ The large size of mAbs restricts their movement through the cell membrane of erythrocytes causing an observable Hct effect; thus, this assay greatly benefits DBS-based mAb studies. A study on home-based use of VAMS by psoriasis patients treated with adalimumab revealed patients preferred VAMS to conventional sampling despite ~50% being uncer-

tain about the reliability of the sampling technique, reflecting the knowledge gap.¹¹⁰

3.3 | Omics-based biomarker research

Microsampling technologies have been evaluated in biomarker research studies for the diagnosis and prognosis of various conditions and the prediction and monitoring of response to an intervention.^{24,41,42,132–144} Omics-based approaches, such as metabolomics and proteomics, are increasingly employed for the generation of large amounts of data to facilitate the rapid discovery of biomarkers. A study assessed the DBS and DPS technology for non-targeted lipidomic analysis as high throughput profiling and quantification of lipids could expand knowledge of underlying disease pathologies. The study's validated assay was capable of annotating 498 compounds covering 24 lipid sub-classes. DPS showed excellent correlation ($R^2 = 0.9851$) to wet plasma, whereas DBS required an Hct correction. However, only 60% of compounds were found to be stable at room temperature after 28 days.¹³² A review of DBS in lipidomics concluded that DBS lipidome by MS/MS analysis could prove to be just as effective as traditional lipidome if DBS stability and extraction are fully optimised taking into account pre-analytical factors such as storage, stabilisers, type of DBS filter paper and Hct.¹³³ The VAMS technology was successfully implemented in a proteomics study to understand the pathophysiology of takotsubo syndrome. The analysis of 10 μL VAMS samples obtained from 80 patients revealed persistent differential protein expression between patients and controls, identifying potential therapeutic targets for recurrence.¹³⁴ VAMS is also a viable microsampling technique for large-scale precision medicine studies using archival frozen blood cell pellets as the detection and robust quantitation of up to 1600 proteins from a single VAMS shotgun analysis demonstrated a higher proteomic depth than a conventional shotgun analysis of undepleted plasma, which only allowed for the quantification of ~324 proteins.¹³⁵ Additionally, VAMS was used to assess the physiological changes associated with the level of training and performance of trail runners which revealed that the fittest trail runners had a better adaptation of bioenergetic pathways.¹³⁶

A novel LC-MS/MS assay measured endogenous uracil levels in TASSO-SST samples to diagnose dihydropyrimidine dehydrogenase deficiency. The study showed no systematic or proportional bias between TASSO-SST serum and venous plasma uracil levels. However, at-home sampling using TASSO-SST for uracil measurements is not feasible due to the compound's high instability. Nevertheless, microsampling technology can be implemented at the hospital to reduce the burden of repeated blood draws in cancer patients.¹³⁷ A randomised, control trial evaluated the accuracy, feasibility and acceptability of microsampling techniques in rheumatoid arthritis patients by measuring inflammation markers and autoantibodies in three matrices: conventional venous blood, upper arm capillary blood obtained using TASSO-SST and finger prick capillary blood. The upper arm and finger prick samples provided comparable results to venous blood samples. Based on a Net Promotor Score, upper arm sampling was found

to be preferred over finger prick sampling by patients.²⁴ Other studies that have implemented the VAMS technology include research for (i) the human biomonitoring of mercury concentrations for exposure assessment,^{41,145} (ii) the determination of tryptophan-related biomarkers for understanding pathologies of neurodegenerative diseases such as Alzheimer's or Parkinson's disease¹³⁸ and (iii) the measurement of endocrine and bone biomarkers.¹³⁹ VAMS technology was also capable of detecting carcinogens, such as acrylamide and aflatoxin B₁.^{140,141}

The incidence of the global COVID-19 pandemic has accelerated the use and development of decentralised blood sampling for virtual clinical trials, trials that do not require in-person visits and routine analysis. The use of TASSO-SST devices was evaluated for measuring blood protein levels in healthy subjects and non-hospitalised COVID-19 patients. The measurements of 17 protein inflammatory biomarkers using a multi-analyte immune-assay revealed that protein concentrations were higher in COVID patients compared to controls. However, a bias between TASSO-SST samples and venous samples was found to differ significantly in D-dimer, IL-1B and IL-1Ra, which could be attributed to differences in user proficiency, temperature control and the time lag between sample collection and processing.⁴² Microsampling has also been used for the serological analysis of SARS-CoV-2 infection in three studies.^{142–144}

3.4 | Forensic toxicology

Forensic toxicology analyses biological samples to determine drug usage, poisoning or exposure to toxic substances for use in a legal investigation. Here, isolation and identification of substances associated with criminal activity has primarily focused on blood samples. Antemortem or post-mortem blood for forensic drug testing or post-mortem toxicology must be appropriately sampled and adequately preserved to ensure the stability of compounds of interest which can have a profound impact on the interpretation of results and the outcome of forensic casework. Additionally, timely collection of biological evidence is a critical component in circumstances where drugs have limited detection windows or collection of post-mortem samples becomes challenging as a result of autolytic and putrefactive changes.¹⁴⁶ Thus, microsampling offers several advantages in forensic toxicological screenings for the analysis of drug abuse: (i) small sample size is beneficial when there is a limited sample, (ii) reduced exposure risk to blood-borne pathogens ensuring the safety of personnel involved in sample collection and analysis and (iii) potential for faster, efficient and economical sample collection and long-term storage with increased stability permitting accurate analysis and re-analysis, even years after sample collection.¹¹⁹

One of the routine forensic toxicology practices is the assessment of alcohol intoxication using the blood alcohol concentration in cases of driving under the influence or traffic incident deaths. However, due to the short half-life of ethanol in blood, typically between 4 and 5 h, other markers are evaluated to determine drinking behaviours and patterns.¹⁴⁷ Phosphatidylethanol (PEth) is a unique

direct ethanol biomarker with a broad detection window to determine alcohol consumption or abstinence. PEth 16:0/18:1, the most prominent homologue in human blood, can act as a complement to the ethyl glucuronide, a biomarker in urine, to distinguish alcohol abusers from social drinkers.¹⁴⁸ In the year 2022, several research articles focused on the analysis of PEth 16:0/18:1 from microsamples acquired via DBS or VAMS technology. Two analytical methods were developed to quantify PEth in DBS samples, where the detection of PEth was achieved using UHPLC-QTOF-HRMS, UHPLC-MS/MS and Orbitrap-based HRMS.^{148,149} In a study quantifying DBS PEth using an LC-MS/MS method, the threshold of PEth for excessive alcohol consumption was determined to be 160 ng/mL when the blood alcohol concentration decision limit was 50 mg/dL. This threshold provided a sensitivity of 92% and a specificity of 90% for predicting excessive alcohol consumption.³⁵ In contrast, a threshold below 20 ng/mL is associated with determining alcohol abstinence. However, given the 4–7 days half-life of PEth, it can take months for PEth values to fall below this threshold following cessation of alcohol consumption. This can make PEth interpretation challenging, particularly when monitoring individuals with alcohol-use disorders. To address this challenge, a population-based algorithm was developed to estimate the cessation of alcohol use based on PEth values acquired from the analysis of VAMS samples obtained at three time points. The decision tree yielded a sensitivity and specificity of 89% by examining the relationship between two successive PEth values that were not yet below the decision limit for alcohol abstinence.³⁶ Although single PEth determinations are beneficial in various forensic contexts, the added value of simultaneously determining three direct ethanol biomarkers was determined: 2 PEth isoforms (PEth 18:1 and PEth 20:4) and ethyl glucuronide in dried blood microsamples by evaluating several forensic cases.¹⁵⁰ Furthermore, it was demonstrated that the distribution of seven different PEth homologues quantified in DBS can aid in better comprehending an individual's drinking history. The study also revealed the risks of sample contamination associated with microsampling for PEth analysis when excessively using ethanol-based hand sanitizers.¹⁵¹

A major challenge in forensic toxicology is the analysis of substances of abuse due to the constant influx of new psychoactive substances (NPS), such as synthetic cannabinoids and opioids, cathinones and hallucinogens, which pose an unprecedented threat to public health. Recently, a UHPLC-HRMS method was developed and validated for the accurate quantification of 132 NPS within an adequate concentration range of 5–100 ng/mL in 30 μ L DBS samples. Most drugs analysed were found to be stable at 4°C for up to 40 days with extremely limited degradation. The method was applied to seven real samples, which provided quantitative results consistent with those obtained using conventional venous whole-blood samples.³⁴ Similarly, cocaine and its metabolites, benzoylecgonine and cocaethylene, were found to be stable for 7 days when DBS samples were retained within the HemaPEN® device.¹⁵² An LC-MS/MS method was also established to detect 425 drugs, including several sedative hypnotics, antileptic drugs and antipsychotic drugs, in 10-mm DBS punch samples. The method was successfully applied to samples obtained from actual drug-poisoning cases that were stored for 3–5 years at room temperature.

The results were in accordance with those measured at the time of sample collection except for a few compounds such as ambroxol, zopiclone, carbofuran and valproic acid, which were not detectable in the long-term preserved DBS samples.¹⁵³ As an alternative to DBS analysis, an HRMS method was developed to screen for 85 licit and illicit drugs in 20 μL VAMS samples, which showed a comparable distribution of substances between capillary and peripheral streams ($R^2 = 0.9997$).¹⁵⁴

Microsampling was also used in the high-throughput identification of toxins and trace elements, where a highly sensitive LC-MS/MS method was developed for the measurement of ricinine and L-abrine in DBS, with respective detection limits of 50 and 100 pg/mL . The method's complete analytical time was less than 1 h, allowing for rapid preliminary confirmation of the presence of highly toxic proteins, ricin and abrin.¹⁵⁵ Although dried blood microsamples have increased stability for a variety of analytes, the authors showed that DBS performed poorly when quantifying trace elements such as aluminium, cadmium, cobalt and chromium, due to high signal contributions from blank filter paper and instability at room temperature. This study demonstrated that microtubes were more suited than DBS for trace element quantification in human blood.¹⁵⁶

3.5 | Sports anti-doping

Microsampling has emerged as a valuable sampling technique for sports anti-doping analysis as it reduces discomfort for athletes and allows for more frequent testing. These potential benefits have also been recognised by the World Anti-Doping Agency (WADA) which has led to the legitimization of DBS for doping control in sports, including the Olympics. Several studies have evaluated the use of microsampling for the detection of prohibited substances, such as anabolic agents, growth factors and hormone and metabolic modulators. A recent critical review by Thevis et al. described the advances, potential future perspectives and limitations of blood microsamples in doping controls.¹⁵⁷ Fast and cost-effective screening methods to detect erythropoietin receptor agonists such as CERA, BRP, NESP and EPO-Fc using dried blood microsamples were developed and optimised.^{37,38} In the study conducted by Heiland et al., different blood sample volumes (20–180 μL) and storage temperatures (-20 to 37°C) were evaluated to determine optimal conditions for detecting prohibited ERAs from DBS. It was discovered that a blood volume of 60 μL allowed the detection of ERAs at levels comparable to WADA's minimum required performance levels (MRPLs) described for 500 μL of serum or plasma.³⁷ In another study, CERA was demonstrated to be detected in DBS using a simple enzyme-linked immunosorbent assay with a screening time of just 2 h.³⁸ Furthermore, a multi-analyte assay was developed and the performance of several non-volumetric (Whatman cards-Treated: Whatman FTA DMPK-A, Whatman FTA DMPK-B; Untreated: Whatman FTA DMPK-C, Whatman 903) and volumetric devices (HemaSpotTM HF, Mitra and TASSO-M20) was evaluated for the simultaneous analysis of 235 drugs for doping control.³⁹ Treated Whatman cards were found to produce very low recoveries or dirty

extracts, whereas all other devices produced high recoveries between 60% and 90%. The LODs of the analytes ranged from 0.1 to 3.0 ng/mL , allowing for the detection of illicit use as most analytes were found to be detectable below the WADA's stipulated MRPL for urine. In addition, a fully automated DBS sample preparation and detection method was also developed for 13 anabolic steroid esters, frequently detected in sports doping controls.¹⁵⁸

4 | Summary and outlook

The advent of microsampling has radically transformed the landscape of blood sampling for various biomedical applications. The continued interest in microsampling has been driven by factors including minimal invasiveness of the approach, decentralised monitoring, logistical benefits, reduced biohazard risk, analytical improvements, data-rich pre-clinical and clinical studies and personalised health monitoring and treatment opportunities. Various microsampling technologies exist for the collection of dried or liquid matrix microsamples. DBS, the oldest microsampling technology, has continuously evolved over the years from a simple, Hct-dependent, non-volumetric collection on filter card to innovative devices for Hct-independent volumetric collection mitigating its limitations of sample heterogeneity and volume bias. Modifications to the inherent filter paper component of DBS have the potential to provide additional functionalities, such as instantaneous proteolysis or stabilization. However, DBS is associated with Hct-based recovery biases that necessitate correction factors, especially for analytes with high plasma-to-RBC partitioning ratio. This bias can be overcome with the use of DPS/DSS which is obtained by filtration of cells through passive separation membranes. This technology inherently eliminates the need for elaborate sample preparation as required in the analysis of whole-blood samples and enables the direct comparison with plasma assays. These 2D dried matrix spots are susceptible to oxidative stress resulting from exposure to ambient air, rendering them unsuitable for the analysis of labile compounds. This vulnerability can be eliminated by utilizing the dried blood spheroid technology, which is a 3D blood collection technique that enhances sample stability by reducing the blood's exposed surface area to air. Alternative to paper-based technologies, VTM produces volumetric dried blood microsamples independent of the Hct using polymeric tips. The dried matrix microsample collection typically involves a finger prick by lancet which enables the collection of 2–80 μL of blood. For collection of larger blood volumes (up to 600 μL), liquid matrix microsamples are collected either using capillary-based or tube-based technologies.

In recent years, the demand for microsampling devices has increased significantly in clinical research and diagnostic applications. This demand has led to a highly competitive market in which manufacturers continuously develop new, innovative and improved devices to offer enhanced performance, greater stability and more user-friendly designs. Currently, the vast majority of commercially-available devices require a simple finger prick by lancet to collect blood onto filter papers, capillaries or volumetric polymer tips. Continuous advance-

ments in lancing device design, such as adjustable depth settings to accommodate varying skin thicknesses and retractable lancets to prevent needle stick injuries, allow for less painful, more efficient and more convenient blood sampling. Self-sampling is particularly beneficial for individuals who need to regularly monitor their health, or who have difficulty with traditional blood collection, such as young children, the elderly or those with medical conditions that make blood collection challenging. Despite the ease of a finger prick in DBS, CMS and VAMS devices, the application of blood droplets onto the pre-printed circle on the filter paper, application surfaces and tips, or their introduction into capillaries, requires manual interventions. These interventions can introduce manual errors and variability, thereby reducing sample quality. Further, CMS and VAMS devices may necessitate a higher level of technical expertise for accurate sampling. In addition, to generate enough volume of blood, lancing is often accompanied by milking of the finger which might result in haemolysis or an increase in the percentage of interstitial fluid collected, affecting downstream measurements. In the future, this problem could potentially be overcome with the use of hollow microneedles to sample blood by generating vacuum. Alternatively, vacuum-assisted technologies to automatically collect blood from upper arm, such as the TASSO and TAP II devices, offer a simple self-sampling design that deploys a lancet/microneedle array and enables automatic blood collection into tubes with the press of a button. In addition to self-sampling capabilities, automation is a desirable feature of microsampling devices because it can significantly enhance the efficiency, accuracy and precision of sample analysis. Several automated DBS sample analysers have been developed to streamline sample analysis and reduce the time and effort required for sample preparation, processing and analysis. These analysers can handle a large number of samples quickly and accurately, making them ideal for high-throughput applications. Similarly, the design of the Mitra samplers is suitable for extraction in microtiter plates, which can readily be further automated with minimal efforts.

Microsampling has been widely employed in different biomedical applications including animal studies, neonatal and paediatric applications, TDM, omics-based biomarker research and forensic toxicology. Microsampling not only reduces the number of animals needed for studies in accordance with the 3Rs principle, but also improves data quality and animal welfare. It reduces the physical and psychological burden of blood sample collection from vulnerable populations such as neonates and young children. With the recent increased focus on personalised healthcare, microsampling provides opportunities for optimizing drug dosages through PK-PD studies and TDM by enabling frequent sampling for data-rich analysis. Longitudinal sampling and sampling during clinical episodes permit the collection of larger datasets for understanding disease biology and identification of diagnostic and prognostic biomarkers. The convenience and accessibility of at-home self-sampling increases patient compliance, diversity and recruitment in clinical trials. In addition, it allows for the secure and efficient collection of samples on-site for forensic toxicology screenings. Several bioanalytical methods have been developed for the accurate quantification of analytes in microsamples using LC-MS/MS analysis.

Despite the profound advantages over conventional sampling, widespread implementation of microsampling in clinical practice has been slow. While several factors contribute to this, a crucial factor is the quality of the samples collected. Implementing at-home self-sampling requires well-defined user instructions on proper sample collection and adherence as well as strategies to monitor patient discipline to ensure sample quality and reduce data variability. Although dried blood microsamples can be shipped at ambient temperatures, extensive evaluation on stability must be performed for the analyte of interest to determine conditions for transportation and storage. Furthermore, optimization of the extraction method is essential in preventing reduced and inconsistent recoveries. Another significant barrier is the perspective of regulatory bodies on patient-centric sampling approaches, which requires bridging studies to be performed to support the introduction of new matrices. Conventional matrices include venous blood, plasma or serum, whereas microsampling provides capillary blood. This inherent difference in sample type and composition can impact concentration of analytes; therefore, method validations and correction factors are necessary to produce reliable data comparable to conventional data. In addition, regular proficiency testing must be conducted to assure the quality, comparability and acceptability of analytical results, as there is no consensus over downstream processing protocols. And, while microsampling minimises the cost of storage, sample transportation and patient travel, the cost-effectiveness must factor in the cost of the microsampling device and its potential limit of being a single-use device.

Microsampling technologies can enable fully automated workflows to provide high-throughput clinical testing. They can also be used in point-of-care devices to either facilitate improved patient care through at-home testing and screenings or aid the development of wearable sensor-enabled technology for continuous health monitoring, leading to digital transformation in healthcare and improved data connectivity. This review demonstrates that considerable advances have been made in recent times; however, data reported for many analytes are inconsistent across microsampling technologies, and in some cases, even for the same technology. Therefore, a continued effort is needed, including standardization, optimization of extraction process, validation of bridging studies and performance evaluation of technologies, for microsampling to be successfully implemented in routine clinical care to reap the benefits of capillary blood microsamples.

AUTHOR CONTRIBUTIONS

Manchu Umarani Thangavelu: Conceptualization (equal); Data curation (lead); Writing – original draft (lead); Writing – review and editing (equal); Bert Wouters: Conceptualization (equal); Writing – review and editing (equal); Alida Kindt: Conceptualization (supporting); Writing – review and editing (equal); Irwin K. M. Reiss: Conceptualization (supporting); Writing – review and editing (supporting); Thomas Han-kemeier: Conceptualization (supporting); Writing – review and editing (supporting).

CONFLICT OF INTEREST STATEMENT

The authors have declared no conflict of interest.

DATA AVAILABILITY STATEMENT

The data that support the findings of this review article are obtained from publicly available sources, including online databases such as PubMed and Google Scholar. Data sharing is not applicable to this article as no new data were created or analysed for the review.

ORCID

Manchu Umarani Thangavelu  <https://orcid.org/0000-0003-2139-3846>

Alida Kindt  <https://orcid.org/0000-0001-6551-6030>

REFERENCES

- Serra R, Ielapi N, Barbetta A, et al. Adverse complications of venipuncture: a systematic review. *Acta Phlebol.* 2018;19:11-15. doi:10.23736/S1593-232X.18.00408-3
- WHO Guidelines on Drawing Blood. World Health Organization; 2010. Accessed January 21, 2023. <https://www.ncbi.nlm.nih.gov/books/NBK138650/>
- Ialongo C, Bernardini S. Phlebotomy, a bridge between laboratory and patient. *Biochem Medica.* 2016;26(1):17-33. doi:10.11613/BM.2016.002
- Buowari OY. Complications of venepuncture. *Adv Biosci Biotechnol.* 2013;4(1):126-128. doi:10.4236/abb.2013.41A018
- Sikorová L, Hrazdilova P. The effect of psychological intervention on perceived pain in children undergoing venipuncture. *Biomed Pap Med Fac Univ Palacký Olomouc Czechoslov.* 2011;155:149-154. doi:10.5507/bp.2011.024
- Bodley T, Chan M, Levi O, et al. Patient harm associated with serial phlebotomy and blood waste in the intensive care unit: a retrospective cohort study. *PLoS One.* 2021;16(1):e0243782. doi:10.1371/journal.pone.0243782
- Bevan EG, Chee LC, McGhee SM, McInnes GT. Patients' attitudes to participation in clinical trials. *Br J Clin Pharmacol.* 1993;35(2):204-207.
- Simundic AM, Lippi G. Preanalytical phase – a continuous challenge for laboratory professionals. *Biochem Medica.* 2012;22(2):145-149.
- Trolio WM, Coville WE. Pre-analytical specimen preparation. *JALA J Assoc Lab Autom.* 2000;5(1):72-78. doi:10.1016/S1535-5535-04-00055-3
- Roberts JL, Whiley L, Gray N, Gay M, Lawler NG. Advanced microsamples: current applications and considerations for mass spectrometry-based metabolic phenotyping pipelines. *Separations.* 2022;9(7):175. doi:10.3390/separations9070175
- Thomas SL, Wakerman J, Humphreys JS. Ensuring equity of access to primary health care in rural and remote Australia - what core services should be locally available? *Int J Equity Health.* 2015;14:111. doi:10.1186/s12939-015-0228-1
- Caron A, Lelong C, Bartels T, et al. Clinical and anatomic pathology effects of serial blood sampling in rat toxicology studies, using conventional or microsampling methods. *Regul Toxicol Pharmacol.* 2015;72(3):429-439. doi:10.1016/j.yrtph.2015.05.022
- Lim MD. Dried blood spots for global health diagnostics and surveillance: opportunities and challenges. *Am J Trop Med Hyg.* 2018;99(2):256-265. doi:10.4269/ajtmh.17-0889
- Lei BUW, Prow T. A review of microsampling techniques and their social impact. *Biomed Microdevices.* 2019;21. doi:10.1007/s10544-019-0412-y
- Londhe V, Rajadhyaksha M. Opportunities and obstacles for microsampling techniques in bioanalysis: special focus on DBS and VAMS. *J Pharm Biomed Anal.* 2020;182:113102. doi:10.1016/j.jpba.2020.113102
- Shahhoseini F, Azizi A, Bottaro CS. Single-use porous polymer thin-film device: a reliable sampler for analysis of drugs in small volumes of biofluids. *Anal Chim Acta.* 2022;1203:339651. doi:10.1016/j.aca.2022.339651
- Seidi S, Tajik M, Baharfar M, Rezazadeh M. Micro solid-phase extraction (pipette tip and spin column) and thin film solid-phase microextraction: miniaturized concepts for chromatographic analysis. *TrAC Trends Anal Chem.* 2019;118:810-827. doi:10.1016/j.trac.2019.06.036
- He Y, Miggiels P, Drouin N, Lindenburg PW, Wouters B, Hankemeier T. An automated online three-phase electro-extraction setup with machine-vision process monitoring hyphenated to LC-MS analysis. *Anal Chim Acta.* 2022;1235:340521. doi:10.1016/j.aca.2022.340521
- Miggiels P, Wouters B, van Westen GJP, Dubbelman AC, Hankemeier T. Novel technologies for metabolomics: more for less. *TrAC Trends Anal Chem.* 2019;120:115323. doi:10.1016/j.trac.2018.11.021
- Zhang W, Ramautar R. CE-MS for metabolomics: developments and applications in the period 2018-2020. *Electrophoresis.* 2021;42(4):381-401. doi:10.1002/elps.202000203
- Sham TT, Badu-Tawiah AK, McWilliam SJ, Maher S. Assessment of creatinine concentration in whole blood spheroids using paper spray ionization-tandem mass spectrometry. *Sci Rep.* 2022;12:14308. doi:10.1038/s41598-022-18365-8
- Biagini D, Antoni S, Ghimenti S, et al. Methodological aspects of dried blood spot sampling for the determination of isoprostanooids and prostanoids. *Microchem J.* 2022;175:107212. doi:10.1016/j.microc.2022.107212
- Ten-Doménech I, Solaz-García Á, Lara-Cantón I, et al. Direct derivatization in dried blood spots for oxidized and reduced glutathione quantification in newborns. *Antioxidants.* 2022;11(6):1165. doi:10.3390/antiox11061165
- Knitz J, Tascilar K, Vuillerme N, et al. Accuracy and tolerability of self-sampling of capillary blood for analysis of inflammation and autoantibodies in rheumatoid arthritis patients-results from a randomized controlled trial. *Arthritis Res Ther.* 2022;24(1):125. doi:10.1186/s13075-022-02809-7
- Carling RS, Emmett EC, Moat SJ. Evaluation of volumetric blood collection devices for the measurement of phenylalanine and tyrosine to monitor patients with phenylketonuria. *Clin Chim Acta.* 2022;535:157-166. doi:10.1016/j.cca.2022.08.005
- Guerra Valero YC, Dorofaeff T, Coulthard MG, et al. Optimal dosing of cefotaxime and desacetylcefotaxime for critically ill paediatric patients. Can we use microsampling? *J Antimicrob Chemother.* 2022;77(8):2227-2237. doi:10.1093/jac/dkac168
- Rosser SPA, Atkinson C, Nath CE, Fletcher JI. Quantification of vincristine and tariquidar by liquid chromatography-tandem mass spectrometry in mouse whole blood using volumetric absorptive microsampling for pharmacokinetic applications. *J Sep Sci.* 2022;45(14):2508-2519. doi:10.1002/jssc.202101013
- Daraghme DN, Moghaddami M, Bobrovskaya L, Proudman SM, Wiese MD. Quantitation of methotrexate polyglutamates in human whole blood, erythrocytes and leukocytes collected via venepuncture and volumetric absorptive micro-sampling: a green LC-MS/MS-based method. *Anal Bioanal Chem.* 2022;414(20):6029-6046. doi:10.1007/s00216-022-04186-1
- Carniel E, dos Santos KA, de Andrade de Lima L, Kohlrausch R, Linden R, Antunes MV. Determination of clozapine and norclozapine in dried plasma spot and dried blood spot by liquid chromatography-tandem mass spectrometry. *J Pharm Biomed Anal.* 2022;210:114591. doi:10.1016/j.jpba.2022.114591
- Parker SL, Wallis SC, Fourie C, et al. Evaluation of low-volume plasma sampling for the analysis of meropenem in clinical samples. *Anal Bioanal Chem.* 2022;414(6):2155-2162. doi:10.1007/s00216-021-03851-1

31. Chiu HH, Tsai YJ, Lo C, et al. Development of an LC-MS/MS method to simultaneously quantify therapeutic mAbs and estimate hematocrit values in dried blood spot samples. *Anal Chim Acta*. 2022;1189:339231. doi:10.1016/j.aca.2021.339231
32. Zwart TC, Metscher E, van der Boog PJM, et al. Volumetric microsampling for simultaneous remote immunosuppressant and kidney function monitoring in outpatient kidney transplant recipients. *Br J Clin Pharmacol*. 2022;88(11):4854-4869. doi:10.1111/bcp.15433
33. Moon SJ, Han SH, Kwak YG, Kim MG. Stability of acetylsalicylic acid in human blood collected using volumetric absorptive microsampling (VAMS) under various drying conditions. *Transl Clin Pharmacol*. 2022;30(1):57-69. doi:10.12793/tcp.2022.30.e5
34. Massano M, Incardona C, Gerace E, et al. Development and validation of a UHPLC-HRMS-QTOF method for the detection of 132 new psychoactive substances and synthetic opioids, including fentanyl, in dried blood spots. *Talanta*. 2022;241:123265. doi:10.1016/j.talanta.2022.123265
35. Daglioglu N, Efeoglu Ozseker P, Dengiz H, Kekec Z. Determination of phosphatidyl ethanol (PEth) 16:0/18:1 in dried blood samples of drivers involved in traffic accidents: a pilot study. *Leg Med Tokyo Jpn*. 2022;58:102091. doi:10.1016/j.legalmed.2022.102091
36. Van Uytanghe K, Heughebaert L, Abatih E, Stove CP. Set-up of a population-based model to verify alcohol abstinence via monitoring of the direct alcohol marker phosphatidylethanol 16:0/18:1. *Addict Abingdon Engl*. 2022;117(7):2108-2118. doi:10.1111/add.15811
37. Heiland CE, Ericsson M, Pohanka A, Ekström L, Marchand A. Optimizing detection of erythropoietin receptor agonists from dried blood spots for anti-doping application. *Drug Test Anal*. 2022;14(8):1377-1386. doi:10.1002/dta.3260
38. Rocca A, Martin L, Kuuranne T, Ericsson M, Marchand A, Leuenberger N. A fast screening method for the detection of CERA in dried blood spots. *Drug Test Anal*. 2022;14(5):820-825. doi:10.1002/dta.3142
39. Mazzarino M, Di Costanzo L, Comunità F, Stacchini C, de la Torre X, Botrè F. UHPLC-HRMS method for the simultaneous screening of 235 drugs in capillary blood for doping control purpose: comparative evaluation of volumetric and non-volumetric dried blood spotting devices. *ACS Omega*. 2022;7(36):31845-31868. doi:10.1021/acsomega.2c01417
40. Radovanovic M, Schneider JJ, Shafiei M, Martin JH, Galettis P. Measurement of 5- fluorouracil, capecitabine and its metabolite concentrations in blood using volumetric absorptive microsampling technology and LC-MS/MS. *J Chromatogr B Analyt Technol Biomed Life Sci*. 2022;1188:123075. doi:10.1016/j.jchromb.2021.123075
41. Koutsimpani-Wagner A, Quartucci C, Rooney JPK, Bose-O'Reilly S, Rakete S. Mercury biomonitoring in German adults using volumetric absorptive microsampling. *Environ Monit Assess*. 2022;194(4):315. doi:10.1007/s10661-022-09962-1
42. Brandsma J, Chenoweth JG, Gregory MK, et al. Assessing the use of a micro-sampling device for measuring blood protein levels in healthy subjects and COVID-19 patients. *PLoS One*. 2022;17(8):e0272572. doi:10.1371/journal.pone.0272572
43. Wouters B, Dapic I, Valkenburg TSE, et al. A cyclic-olefin-copolymer microfluidic immobilized-enzyme reactor for rapid digestion of proteins from dried blood spots. *J Chromatogr A*. 2017;1491:36-42. doi:10.1016/j.chroma.2017.01.078
44. Hallworth MJ. The '70% claim': what is the evidence base? *Ann Clin Biochem*. 2011;48(6):487-488. doi:10.1258/acb.2011.011177
45. At-home blood collection and micro sampling devices market by type of blood sample state, method of sample collection, device usage, area of application, and key geographical regions - industry trends and global forecasts, 2021-2030. *Research and Markets*. Accessed September 16, 2022. <https://www.researchandmarkets.com/reports/5470966/at-home-blood-collection-and-micro-sampling>
46. Bang IC. *Der Blutzucker*. Wiesbaden, Bergmann; 1913. Accessed April 24, 2023. <http://archive.org/details/derblutzucker00bang>
47. American College of Medical Genetics Newborn Screening Expert Group. Newborn screening: toward a uniform screening panel and system-executive summary. *Pediatrics*. 2006;117(5 Pt 2):S296-307. doi:10.1542/peds.2005-2633I
48. Enderle Y, Foerster K, Burhenne J. Clinical feasibility of dried blood spots: analytics, validation, and applications. *J Pharm Biomed Anal*. 2016;130:231-243. doi:10.1016/j.jpba.2016.06.026
49. De Kesel PM, Sadones N, Capiou S, Lambert WE, Stove CP. Hematocritical issues in quantitative analysis of dried blood spots: challenges and solutions. *Bioanalysis*. 2013;5(16):2023-2041. doi:10.4155/bio.13.156
50. Wilhelm AJ, den Burger JCG, Swart EL. Therapeutic drug monitoring by dried blood spot: progress to date and future directions. *Clin Pharmacokinet*. 2014;53(11):961-973. doi:10.1007/s40262-014-0177-7
51. Hall E, Flores S, De Jesus V. Influence of hematocrit and total-spot volume on performance characteristics of dried blood spots for newborn screening. *Int J Neonatal Screen*. 2015;1:69-78. doi:10.3390/ijns1020069
52. Lenk G, Hansson J, Beck O, Roxhed N. The effect of drying on the homogeneity of DBS. *Bioanalysis*. 2015;7(16):1977-1985. doi:10.4155/bio.15.135
53. Klak A, Pauwels S, Vermeersch P. Preanalytical considerations in therapeutic drug monitoring of immunosuppressants with dried blood spots. *Diagn Berl Ger*. 2019;6(1):57-68. doi:10.1515/dx-2018-0034
54. Zadow JG, Hill RD. The precipitation of proteins by carboxymethyl cellulose. *J Dairy Res*. 1975;42(2):267-275. doi:10.1017/s0022029900015302
55. Mengerink Y, Mommers J, Qiu J, Mengerink J, Steijger O, Honing M. A new DBS card with spot sizes independent of the hematocrit value of blood. *Bioanalysis*. 2015;7(16):2095-2104. doi:10.4155/bio.15.133
56. Baillargeon KR, Brooks JC, Miljanic PR, Mace CR. Patterned dried blood spot cards for the improved sampling of whole blood. *ACS Meas Sci Au*. 2022;2(1):31-38. doi:10.1021/acsmesuresciau.1c00031
57. HemaSpot-HD. Product Summary Sheet. Spot On Sciences. September, 2022. Accessed April 18, 2023. <https://www.spotonsciences.com/wp-content/uploads/HemaSpot-HD-Product-Brochure-Rev-A-Sep-2022.pdf>
58. HemaXis DB 10. Instructions for use. HemaXis. Published December 17, 2017. Accessed April 24, 2023. <http://hemaxis.com/wp-content/uploads/2018/12/HemaXis-DB-10-IFU-EN-FR.pdf>
59. Capitainer®B - Quantitative dried blood spot sampling. Capitainer. Accessed April 24, 2023. <https://capitainer.se/capitainerb/>
60. HemaPEN. Advanced precision microsampling. Trajan Scientific and Medical. Published March, 2021. Accessed April 24, 2023. https://cdn.shopify.com/s/files/1/0767/9441/files/BR-0536-G_RevD.pdf?v=1618268327
61. HemaSpot-HF. Product Summary Sheet. Spot On Sciences. September, 2022. Accessed April 24, 2023. <https://www.spotonsciences.com/wp-content/uploads/HemaSpot-HF-Product-Brochure-Rev-A-Sep-2022.pdf>
62. Denniff P, Spooner N. The effect of hematocrit on assay bias when using DBS samples for the quantitative bioanalysis of drugs. *Bioanalysis*. 2010;2(8):1385-1395. doi:10.4155/bio.10.103
63. Mathew BS, Mathew SK, Aruldas BW, et al. Analytical and clinical validation of dried blood spot and volumetric absorptive microsampling for measurement of tacrolimus and creatinine after renal transplantation. *Clin Biochem*. 2022;105-106:25-34. doi:10.1016/j.clinbiochem.2022.04.014
64. Paniagua-González L, Lendoiro E, Otero-Antón E, López-Rivadulla M, de-Castro-Ríos A, Cruz A. Comparison of conventional dried blood spots and volumetric absorptive microsampling for tacrolimus and mycophenolic acid determination. *J Pharm Biomed Anal*. 2022;208:114443. doi:10.1016/j.jpba.2021.114443

65. Xiaoyong X, Jinglin W, Guangfei W, Huimin Z, Hong X, Zhiping L. *Applicability of Vancomycin, Meropenem and Linezolid in Capillary Microsamples versus Dried Blood Spots: A Validation and Pilot Study Suitable for Microsampling in Critically Ill Children*. Review;2022. doi:10.21203/rs.3.rs-2010514/v1
66. Luo M, Chen Y, Cheng Y, Li N, Qing H. Association between hematocrit and the 30-day mortality of patients with sepsis: a retrospective analysis based on the large-scale clinical database MIMIC-IV. *PLoS One*. 2022;17(3):e0265758. doi:10.1371/journal.pone.0265758
67. Nguyen M, Halvorsen T, Thiede B, Reubsaet L. Back cover: smart proteolysis samplers for pre-lab bottom-up protein analysis – performance of on-paper digestion compared to conventional digestion. *Sep Sci PLUS*. 2022;5. doi:10.1002/sscp.202270036
68. Mercatali L, Serra P, Miserocchi G, et al. Dried blood and serum spots as a useful tool for sample storage to evaluate cancer biomarkers. *J Vis Exp JoVE*. 2018;136:57113. doi:10.3791/57113
69. Kim JH, Woenker T, Adamec J, Regnier FE. Simple, Miniaturized blood plasma extraction method. *Anal Chem*. 2013;85(23):11501-11508. doi:10.1021/ac402735y
70. HemaSpot-SE. Product Summary Sheet. Spot on Sciences. November 2022. Accessed April 24, 2023. <https://www.spotonsciences.com/wp-content/uploads/HemaSpot-SE-Product-Brochure-Rev-C.-Nov-2022.pdf>
71. Frey BS, Damon DE, Allen DM, Baker J, Asamoah S, Badu-Tawiah AK. Protective mechanism of dried blood spheroids: stabilization of labile analytes in whole blood, plasma, and serum. *Analyst*. 2021;146(22):6780-6787. doi:10.1039/d1an01132d
72. Damon DE, Yin M, Allen DM, et al. Dried blood spheroids for dry-state room temperature stabilization of microliter blood samples. *Anal Chem*. Published online July 5, 2018. doi:10.1021/acs.analchem.8b01962
73. Damon DE, Davis KM, Moreira CR, Capone P, Cruttenden R, Badu-Tawiah AK. *Direct Biofluid Analysis Using Hydrophobic Paper Spray Mass Spectrometry*. ACS Publications. doi:10.1021/acs.analchem.5b04278
74. Frey BS, Heiss DR, Badu-Tawiah AK. Embossed-paper platform for whole blood collection, room temperature storage, and direct analysis by pinhole paper spray mass spectrometry. *Anal Chem*. 2022;94(10):4417-4425. doi:10.1021/acs.analchem.1c05340
75. Mitra Cartridge | DBS Collection | Capillary Blood Sampling Device. *Trajan Scientific and Medical*. 2022. Accessed January 31, 2023. <https://www.neoteryx.com/mitra-cartridge-blood-sampling-device-dbs>
76. TASSO-M20. Tasso, Inc. Accessed October 9, 2022. <https://www.tasso-inc.com/tasso-m20>
77. Kok MGM, Fillet M. Volumetric absorptive microsampling: current advances and applications. *J Pharm Biomed Anal*. 2018;147:288-296. doi:10.1016/j.jpba.2017.07.029
78. Verougstraete N, Stove C. Volumetric absorptive microsampling as a suitable tool to monitor tyrosine kinase inhibitors. *J Pharm Biomed Anal*. 2022;207. doi:10.1016/j.jpba.2021.114418
79. Reubsaet L, Thiede B, Halvorsen TG. Next generation VAMS®-Trypsin immobilization for instant proteolysis in bottom-up protein determination. *Adv Sample Prep*. 2022;3:100027. doi:10.1016/j.sampre.2022.100027
80. Jonsson O. Capillary microsampling. In: Zane P, Emmons GT, eds. *Microsampling in Pharmaceutical Bioanalysis*. Future Science Book Series. Future Science Ltd; 2013:68-82. doi:10.4155/ebo.13.399
81. Nys G, Kok MGM, Servais AC, Fillet M. Beyond dried blood spot: current microsampling techniques in the context of biomedical applications. *TrAC Trends Anal Chem*. 2017;97:326-332. doi:10.1016/j.trac.2017.10.002
82. Minivette POCT. Collection and dispensing of small blood samples for Point-of-Care tests. *Sarstedt*. Accessed April 24, 2023. https://www.sarstedt.com/fileadmin/user_upload/99_Broschueren/Broschueren_neu_05.2016/481_g_Minivette_EN_0716_www.pdf
83. Microvette APT. Automated processing tube for routine capillary blood analysis. *Sarstedt*. Accessed April 13, 2023. https://www.sarstedt.com/fileadmin/user_upload/99_Broschueren/NEU/865/20_865_0000_200_microvette_apt_1020.pdf
84. MSW² Type Udck. *Microsampling Device*. Shimadzu Europa. Accessed December 4, 2022. <https://www.shimadzu.fr/msw2>
85. TASSO+. Tasso, Inc. Accessed October 16, 2022. <https://www.tasso-inc.com/tasso-plus>
86. TASSO-SST. Tasso, Inc. Accessed October 16, 2022. <https://www.tasso-inc.com/tasso-sst>
87. TAP II Product Overview. YourBio Health, Inc. Accessed October 16, 2022. <https://yourbiohealth.com/tap-ii-product-overview>
88. El-Laboudi A, Oliver NS, Cass A, Johnston D. Use of microneedle array devices for continuous glucose monitoring: a review. *Diabetes Technol Ther*. 2013;15(1):101-115. doi:10.1089/dia.2012.0188
89. The 3Rs. National Centre for the Replacement, Refinement & Reduction of Animals in Research. Accessed November 21, 2022. <https://www.nc3rs.org.uk/who-we-are/3rs>
90. Chapman K, Chivers S, Gliddon D, et al. Overcoming the barriers to the uptake of nonclinical microsampling in regulatory safety studies. *Drug Discov Today*. 2014;19(5):528-532. doi:10.1016/j.drudis.2014.01.002
91. Londhe V, Rajadhyaksha M. Pertinence of microsampling in various nonclinical and clinical studies: the ICH perspective. *Bioanalysis*. Published online September 16, 2022. doi:10.4155/bio-2022-0163
92. Xu Y, Zhang S, Xia G, Guo J, et al. A joint technology combining the advantages of capillary microsampling with mass spectrometry applied to the trans-resveratrol pharmacokinetic study in mice. *J Anal Methods Chem*. 2022;2022:e5952436. doi:10.1155/2022/5952436
93. Ryu S, Hayashi Y, Yagishita S, et al. Development of an analytical method to determine E7130 concentration in mouse plasma by micro-sampling using ultra-performance liquid chromatography–high resolution mass spectrometry. *J Chromatogr B*. 2022;1207:123366. doi:10.1016/j.jchromb.2022.123366
94. Pelcová M, Krejčířová E, Juřica J, Glatz Z. Pharmacokinetic profiling via dried blood spot sampling method. Published online 2022. Accessed November 22, 2022. <https://www.muni.cz/en/research/publications/2212599>
95. Verhaeghe T, Dillen L, Stieltjes H, et al. The application of capillary microsampling in GLP toxicology studies. *Bioanalysis*. 2017;9(7):531-540. doi:10.4155/bio-2016-0297
96. Tochitani T, Sasaki Y, Nishimura N, et al. Effects of microsampling on toxicity assessment of hematotoxic compounds in a general toxicity study in rats. *J Toxicol Sci*. 2022;47(7):269-276. doi:10.2131/jts.47.269
97. Manuppello J, Slankster-Schmierer E, Baker E, Sullivan K. Animal use and opportunities for reduction in carcinogenicity studies supporting approved new drug applications in the U.S., 2015-2019. *Regul Toxicol Pharmacol*. 2023;137:105289. doi:10.1016/j.yrtph.2022.105289
98. Grunau RE, Holsti L, Peters JWB. Long-term consequences of pain in human neonates. *Semin Fetal Neonatal Med*. 2006;11(4):268-275. doi:10.1016/j.siny.2006.02.007
99. Goto T, Inoue T, Kamiya C, et al. Neonatal pain response to automatic lancet versus needle heel-prick blood sampling: a prospective randomized controlled clinical trial. *Pediatr Int*. 2020;62(3):357-362. doi:10.1111/ped.14142
100. Anibaletto dos Santos AL, Cezimbra da Silva AC, de Feltraco Lizot L L, et al. Development and validation of an assay for the measurement of gentamicin concentrations in dried blood spots using UHPLC-MS/MS. *J Pharm Biomed Anal*. 2022;208:114448. doi:10.1016/j.jpba.2021.114448
101. Parker SL, Irwin AD, Hosking F, et al. Innovation in microsampling for therapeutic drug monitoring of gentamicin in neonates: a proof-of-concept study. *Int J Antimicrob Agents*. 2022;59(2):106513. doi:10.1016/j.ijantimicag.2021.106513

102. Huang L, Winger BA, Cheah V, et al. Quantification of N, N' N"-triethylenephosphoramidate, N, N' N"-triethylenephosphoramidate, cyclophosphamide, and 4-hydroxy-cyclophosphamide in microvolume human plasma to support neonatal and pediatric drug studies. *J Chromatogr Open*. 2022;2:100054. doi:10.1016/j.jcoa.2022.100054
103. Bamat NA, Vedar C, Reilly ME, Moorthy GS, Zuppa AF. A whole blood microsampling furosemide assay: development, validation and use in a pediatric pharmacokinetic study. *Bioanalysis*. 2022;14(15):1025-1038. doi:10.4155/bio-2022-0063
104. Simeoli R, Dorlo TPC, Hanff LM, Huitema ADR, Dreesen E. Editorial: therapeutic drug monitoring (TDM): a useful tool for pediatric pharmacology applied to routine clinical practice. *Front Pharmacol*. 2022;13. Accessed November 29, 2022. <https://www.frontiersin.org/articles/10.3389/fphar.2022.931843>
105. Bachhav SS, Taylor M, Martin A, et al. A pharmacometrics approach to assess the feasibility of capillary microsampling to replace venous sampling in clinical studies: Tafenoquine case study. *Br J Clin Pharmacol*. Published online October 5, 2022. doi:10.1111/bcp.15554
106. Guerra Valero Y, Dorofaeff T, Parker L, et al. Microsampling to support pharmacokinetic clinical studies in pediatrics. *Pediatr Res*. 2022;91(6):1557-1561. doi:10.1038/s41390-021-01586-4
107. Zhuang YJ, Mangwiro Y, Wake M, Saffery R, Greaves RF. Multi-omics analysis from archival neonatal dried blood spots: limitations and opportunities. *Clin Chem Lab Med CCLM*. 2022;60(9):1318-1341. doi:10.1515/cclm-2022-0311
108. Anand V, Pournami F, Prithvi AK, Nandakumar A, Prabhakar J, Jain N. Every treasured drop! Blood transfusion requirements in very preterm neonates after implementation of blood conservation strategies: an observational analytical study. *J Trop Pediatr*. 2022;68(6):fmac093. doi:10.1093/tropej/fmac093
109. Vogenberg FR, Isaacson Barash C, Pursel M. Personalized medicine: part 1: evolution and development into theranostics. *P T Peer-Rev J Formul Manag*. 2010;35:560-576.
110. Soenen R, Stove C, Capobianco A, et al. Promising tools to facilitate the implementation of TDM of biologics in clinical practice. *J Clin Med*. 2022;11(11):3011. doi:10.3390/jcm11113011
111. Tey HY, See HH. A review of recent advances in microsampling techniques of biological fluids for therapeutic drug monitoring. *J Chromatogr A*. 2021;1635:461731. doi:10.1016/j.chroma.2020.461731
112. Bressán IG, Giménez MI, Llesuy SF. Clinical validation of a liquid chromatography-tandem mass spectrometry method for the quantification of calcineurin and mTOR inhibitors in dried matrix on paper discs. *J Mass Spectrom Adv Clin Lab*. 2022;25:12-18. doi:10.1016/j.jmsacl.2022.06.002
113. Leino AD, Takyi-Williams J, Wen B, Sun D, Pai MP. Application of a new volumetric microsampling device for quantitative bioanalysis of immunosuppression. *Bioanalysis*. Published online October 8, 2022. doi:10.4155/bio-2022-0155
114. Deprez S, Heughebaert L, Boffel L, Stove CP. Application of non-contact hematocrit prediction technologies to overcome hematocrit effects on immunosuppressant quantification from dried blood spots. *Talanta*. 2023;254:124111. doi:10.1016/j.talanta.2022.124111
115. Veenhof H, Koster RA, Junier LAT, Zweipfening P, Touw DJ. Results from a proficiency testing pilot for immunosuppressant microsampling assays. *Ther Drug Monit*. Published online April 4, 2022: 10.1097/FTD.0000000000001019
116. Deprez S, Stove CP. Dried blood microsampling-assisted therapeutic drug monitoring of immunosuppressants: an overview. *J Chromatogr A*. 2023;1689:463724. doi:10.1016/j.chroma.2022.463724
117. Cao H, Li L, Wang S, et al. Dried plasma spot-based liquid chromatography-tandem mass spectrometry for the quantification of methotrexate in human plasma and its application in therapeutic drug monitoring. *J Sep Sci*. 2022;45(6):1153-1161. doi:10.1002/jssc.202100700
118. Tang Y, Chen L, Liang X, Dean B, Wang J. Exploring the potential of dried plasma collection cards for liquid chromatography coupled with tandem mass spectrometry quantitation of giredestrant in human plasma. *Biomed Chromatogr*. 2023;37:e5554. doi:10.1002/bmc.5554
119. Harahap Y, Steven S, Suryadi H. Development and validation of a UPLC-MS/MS method with volumetric absorptive microsampling to quantitate cyclophosphamide and 4-hydroxycyclophosphamide. *Front Pharmacol*. 2022;13:928721. doi:10.3389/fphar.2022.928721
120. Opitz P, Zimmermann S, Mc Laughlin AM, et al. Development and validation of a bioanalytical method for the quantification of axitinib from plasma and capillary blood using volumetric absorptive microsampling (VAMS) and on-line solid phase extraction (SPE) LC-MS. *J Pharm Biomed Anal*. 2022;221:115033. doi:10.1016/j.jpba.2022.115033
121. Zimmermann S, Aghai F, Schilling B, et al. Volumetric absorptive microsampling (VAMS) for the quantification of ten kinase inhibitors and determination of their in vitro VAMS-to-plasma ratio. *J Pharm Biomed Anal*. 2022;211:114623. doi:10.1016/j.jpba.2022.114623
122. Arora A, Scholar EM. Role of tyrosine kinase inhibitors in cancer therapy. *J Pharmacol Exp Ther*. 2005;315(3):971-979. doi:10.1124/jpet.105.084145
123. Ergun Y, Yildirim Ozdemir N, Toptas S, et al. Drug-drug interactions in patients using tyrosine kinase inhibitors: a multicenter retrospective study. *J BUON Off J Balk Union Oncol*. 2019;24(4):1719-1726.
124. Verougstraete N, Stove V, Verstraete AG, Stove CP. Therapeutic drug monitoring of tyrosine kinase inhibitors using dried blood microsamples. *Front Oncol*. 2022;12. Accessed October 13, 2022. <https://www.frontiersin.org/articles/10.3389/fonc.2022.821807>
125. Krützmann ME, Martini RR, de Souza Guterres F, et al. Volumetric dried blood microsampling for monitoring imatinib mesylate therapy: method development and clinical application in patients with chronic myeloid leukemia. *J Pharm Biomed Anal*. 2023;222:115108. doi:10.1016/j.jpba.2022.115108
126. Marasca C, Mandrioli R, Sardella R, et al. Dried volumetric microsampling approaches for the therapeutic drug monitoring of psychiatric patients undergoing clozapine treatment. *Front Psychiatry*. 2022;13:794609. doi:10.3389/fpsy.2022.794609
127. Jacobs C, Kunz M, Waggmann L, Mahfoud F, Meyer M. Adherence monitoring of antihypertensive drugs in finger prick blood microsamples – development and validation of an analytical procedure. *Toxicol Anal Clin*. 2022;34(3, Supplement):S41. doi:10.1016/j.toxac.2022.06.042
128. Boffel L, Delahaye L, De Baerdemaeker L, Stove CP. Application of a volumetric absorptive microsampling (VAMS)-based method for the determination of paracetamol and four of its metabolites as a tool for pharmacokinetic studies in obese and non-obese patients. *Clin Pharmacokinet*. 2022;61(12):1719-1733. doi:10.1007/s40262-022-01187-2
129. Azzahra Rahmadhani C, Harahap Y, Aisyah Rahmanian T. HPLC-DAD quantification of favipiravir in whole blood after extraction from volumetric absorptive microsampling devices. *J Chromatogr B*. 2023;1215:123547. doi:10.1016/j.jchromb.2022.123547
130. Lo Faro AF, Tini A, Gottardi J, et al. UHPLC-MS-MS determination of THC, CBD and their metabolites in whole blood of light cannabis smokers. *J Anal Toxicol*. Published online October 10, 2022:bkac081. doi:10.1093/jat/bkac081
131. Pigliasco F, Malaca S, Lo Faro AF, et al. Cannabidiol, Δ^9 -tetrahydrocannabinol, and metabolites in human blood by volumetric absorptive microsampling and LC-MS/MS following controlled administration in epilepsy patients. *Front Pharmacol*. 2022;13:1038754. doi:10.3389/fphar.2022.1038754
132. Bishop LM, Fiehn O. Comprehensive lipidomic profiling by plasma separation cards. *Anal Bioanal Chem*. Published online November 1, 2022. doi:10.1007/s00216-022-04399-4

133. Ferreira HB, Guerra IMS, Melo T, et al. Dried blood spots in clinical lipidomics: optimization and recent findings. *Anal Bioanal Chem.* 2022;414(24):7085-7101. doi:[10.1007/s00216-022-04221-1](https://doi.org/10.1007/s00216-022-04221-1)
134. Washington K, Marano P, Chazarin B, et al. Remote microsample blood collection and proteomic analysis of patients with prior takotsubo cardiomyopathy shows profile distinct from normal controls. *J Am Coll Cardiol.* 2022;79(9_Supplement):2033-2033. doi:[10.1016/S0735-1097\(22\)03024-8](https://doi.org/10.1016/S0735-1097(22)03024-8)
135. Molloy MP, Hill C, O'Rourke MB, et al. Proteomic analysis of whole blood using volumetric absorptive microsampling for precision medicine biomarker studies. *J Proteome Res.* 2022;21(4):1196-1203. doi:[10.1021/acs.jproteome.1c00971](https://doi.org/10.1021/acs.jproteome.1c00971)
136. Puigarnau S, Fernández A, Obis E, et al. Metabolomics reveals that fittest trail runners show a better adaptation of bioenergetic pathways. *J Sci Med Sport.* 2022;25(5):425-431. doi:[10.1016/j.jsams.2021.12.006](https://doi.org/10.1016/j.jsams.2021.12.006)
137. Menestrina Dewes M, Cé da Silva L, Fazenda Meireles Y, et al. Evaluation of the Tasso-SST® capillary blood microsampling device for the measurement of endogenous uracil levels. *Clin Biochem.* 2022;107:1-6. doi:[10.1016/j.clinbiochem.2022.06.003](https://doi.org/10.1016/j.clinbiochem.2022.06.003)
138. Protti M, Cirrincione M, Mandrioli R, et al. Volumetric absorptive microsampling (VAMS) for targeted LC-MS/MS determination of tryptophan-related biomarkers. *Molecules.* 2022;27(17):5652. doi:[10.3390/molecules27175652](https://doi.org/10.3390/molecules27175652)
139. Dunn RF, Washbourne CJ, Greeves J, Fraser WD, Tang JCY. Is finger prick blood collection using Mitra® volumetric absorptive microsampling (VAMS) device a viable alternative to venous for testosterone, cortisol, 25 hydroxyvitamin D and bone resorption marker [beta]-CTX measurements? In: *Endocrine Abstracts*. Vol 86. Bioscientifica; 2022. doi:[10.1530/endoabs.86.P28](https://doi.org/10.1530/endoabs.86.P28)
140. Ikhsan M, Harahap Y, Saputri FC. development and validation of HPLC-UV method for simultaneous analysis of acrylamide and glycidamide in volumetric absorptive microsampling. *Int J Appl Pharm.* 2022;14(5):170-174. doi:[10.22159/ijap.2022v14i5.45440](https://doi.org/10.22159/ijap.2022v14i5.45440)
141. Renaud JB, Walsh JP, Sumarah MW. Optimization of aflatoxin B1-lysine analysis for public health exposure studies. *Toxins.* 2022;14(10):672. doi:[10.3390/toxins14100672](https://doi.org/10.3390/toxins14100672)
142. McCarthy P, Pathakamuri JA, Kuebler D, et al. A novel dry-stabilized whole blood microsampling and protein extraction method for testing of SARS-CoV-2 antibody titers. *Vaccines.* 2022;10(10):1760. doi:[10.3390/vaccines10101760](https://doi.org/10.3390/vaccines10101760)
143. Michielin G, Arefi F, Puhach O, et al. Clinical sensitivity and specificity of a high-throughput microfluidic nano-immunoassay combined with capillary blood microsampling for the identification of anti-SARS-CoV-2 Spike IgG serostatus. Published online June 14, 2022;2022.06.09.22276142. doi:[10.1101/2022.06.09.22276142](https://doi.org/10.1101/2022.06.09.22276142)
144. Gordon-Lipkin EM, Marcum C, Kruk S, et al. Undiagnosed COVID-19 in households with a child with mitochondrial disease. *MedRxiv Prepr Serv Health Sci.* Published online March 23, 2022;2022.03.21.22272358. doi:[10.1101/2022.03.21.22272358](https://doi.org/10.1101/2022.03.21.22272358)
145. Gonzalez Rubio JM, González-Rubio JM, Domínguez-Morueco N, et al. Applicability of dried blood spot (Dbs) samples for assessing mercury exposure in human biomonitoring studies. Published online December 20, 2022. doi:[10.2139/ssrn.4308049](https://doi.org/10.2139/ssrn.4308049)
146. Hibou V. Blood Sampling » Analytical Toxicology. Analytical Toxicology. Published November 25, 2016. Accessed November 19, 2022. <https://www.analyticaltoxicology.com/en/blood-sampling/>
147. Van Uytanghe K, De Boosere E, Stove CP. Monitoring the use of alcohol—a critical overview of the state-of-the-art biomarkers. *WIREs Forensic Sci.* 2022;4(5):e1457. doi:[10.1002/wfs2.1457](https://doi.org/10.1002/wfs2.1457)
148. Ververi C, Massano M, Gerace E, Alladio E, Vincenti M, Salomone A. Phosphatidylethanol (Peth) in dried blood spots: development, validation and comparison between LC-MS/MS and QTOF methods. *Toxicol Anal Clin.* 2022;34(3, Supplement):S54. doi:[10.1016/j.toxac.2022.06.065](https://doi.org/10.1016/j.toxac.2022.06.065)
149. Déglon J, Joye T, Lauer E, Thomas A, Augsburger M. Quantification of phosphatidylethanol combining dried blood spots sampling with Orbitrap-based LC-HRMS platform: Method validation and proficiency testing. *Toxicol Anal Clin.* 2022;34(3, Supplement):S106. doi:[10.1016/j.toxac.2022.06.166](https://doi.org/10.1016/j.toxac.2022.06.166)
150. Hakim F, Ghoul C, Wiart JF, et al. Added value evidences of a simultaneous determination of two phosphatidylethanol isoforms and ethylglucuronide in dried blood spots in forensic contexts. *Toxicol Anal Clin.* 2022;34(3, Supplement):S36. doi:[10.1016/j.toxac.2022.06.033](https://doi.org/10.1016/j.toxac.2022.06.033)
151. Herzog J, Skopp G, Musshoff F. Development and validation of seven phosphatidylethanol-homologues in dried blood spots including preliminary results after excessive use of an ethanol-based hand sanitizer. *J Anal Toxicol.* Published online October 26, 2022;bkac086. doi:[10.1093/jat/bkac086](https://doi.org/10.1093/jat/bkac086)
152. Smidt M, Bastiani MF, Hahn RZ, Lima Feltraco Lizot L de, Perassolo MS, Linden R. Evaluation of hemaPEN® sampling device for measurement of cocaine and metabolites in capillary blood by LC-MS/MS. *Bioanalysis.* 2022;14(20):1295-1303. doi:[10.4155/bio-2022-0192](https://doi.org/10.4155/bio-2022-0192)
153. Guo C, Yan H, Liu W, Xiang P, Di B, Shen M. Development of an Lc-MS/Ms method for determining 425 drugs in dried blood spots and application to forensic cases. Published online May 2, 2022. doi:[10.2139/ssrn.4098498](https://doi.org/10.2139/ssrn.4098498)
154. Houzé P, Borowski I, Bito E, Megarbane B, Labat L. Micro sampling, a new trend in toxicological screening? *Toxicol Anal Clin.* 2022;34(3, Supplement):S50-S51. doi:[10.1016/j.toxac.2022.06.059](https://doi.org/10.1016/j.toxac.2022.06.059)
155. Yishai Aviram L, Loewenthal D, Hindi A, Gura S, Weissberg A, Dagan S. Dried urine spot and dried blood spot sample collection for rapid and sensitive monitoring of exposure to ricin and abrin by LC-MS/MS analysis of ricinine and l-abrine. *Forensic Chem.* 2022;30:100438. doi:[10.1016/j.forc.2022.100438](https://doi.org/10.1016/j.forc.2022.100438)
156. Perrais M, Thomas A, Augsburger M, Lenglet S. Comparison of dried blood spot and microtube techniques for trace element quantification by ICP-MS. *J Anal Toxicol.* Published online August 5, 2022;bkac054. doi:[10.1093/jat/bkac054](https://doi.org/10.1093/jat/bkac054)
157. Thevis M, Walpurgis K, Thomas A. DropWise: current role and future perspectives of dried blood spots (DBS), blood microsampling, and their analysis in sports drug testing. *Crit Rev Clin Lab Sci.* 2022;0(0):1-22. doi:[10.1080/10408363.2022.2103085](https://doi.org/10.1080/10408363.2022.2103085)
158. Jing J, Shan Y, Liu Z, et al. Automated online dried blood spot sample preparation and detection of anabolic steroid esters for sports drug testing. *Drug Test Anal.* 2022;14(6):1040-1052. doi:[10.1002/dta.3226](https://doi.org/10.1002/dta.3226)

How to cite this article: Thangavelu MU, Wouters B, Kindt A, Reiss IKM, Hankemeier T. Blood microsampling technologies: Innovations and applications in 2022. *Anal Sci Adv.* 2023;1-27. <https://doi.org/10.1002/ansa.202300011>